

ISO 27186:2010-03 (E)

Active implantable medical devices - Four-pole connector system for implantable cardiac rhythm management devices - Dimensional and test requirements

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	5
Annex A (normative) Electrical isolation test	22
Annex B (informative) Rationale for Annex A	27
Annex C (normative) Dielectric strength test	29
Annex D (informative) Rationale for Annex C	35
Annex E (normative) Current-carrying test high voltage types	39
Annex F (informative) Rationale for Annex E	44
Annex G (informative) Lead connector fatigue strength test	46
Annex H (informative) Lead connector seal zone materials	47
Annex I (informative) Seal zone creep	49
Annex J (informative) Contact resistance stability	54
Annex K (informative) Rationale for Annex J	58
Annex L (informative) Selection of contact materials	60
Annex M (normative) Lead connector contact material requirements	62
Annex N (informative) Rationale for Annex M	66
Annex O (informative) Rationale for requirements in this International Standard	72
Annex P (informative) Connector products (e.g. adaptors, extenders, patient cables, etc.)	79
Bibliography	81