

ISO 80601-2-55:2011-12 (E)

Medical electrical equipment - Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors

Contents	Page
Foreword	vi
Introduction.....	vii
1 Scope	1
201.1 Scope, object and related standards.....	1
201.1. 1 * Scope	1
201.1. 2 Object.....	2
201.1. 3 Collateral standards	2
201.1. 4 Particular standards	2
201.2 Normative references	3
201.3 Terms and definitions	4
201.4 General requirements	6
201.4. 3 ESSENTIAL PERFORMANCE	6
201.4. 3.101 * Additional requirements for ESSENTIAL PERFORMANCE	6
201.4. 3.102 Additional requirements for acceptance criteria	6
201.4. 6 * ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT	6
201.4.10.2.101 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	6
201.5 General requirements for testing ME EQUIPMENT	7
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	7
201.7 ME EQUIPMENT identification, marking and documents	7
201.7. 2.3 * Consult ACCOMPANYING DOCUMENTS.....	7
201.7. 2.101 * Additional requirements for marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts	7
201.7. 2.4.101 Additional requirements for ACCESSORIES	8
201.7. 2.13.101 * Additional requirements for physiological effects (safety signs and warning statements)	8
201.7. 2.17.101 Additional requirements for protective packaging	8
201.7. 4.3 Unit of measure	8
201.7. 9.1 General requirements	9
201.7. 9.2.1.101 * Additional general requirements.....	9
201.7. 9.2.2.101 * Additional requirements for warnings and safety notices	9
201.7. 9.2.5.101 Additional requirements for ME EQUIPMENT description	10
201.7. 9.2.8.101 * Additional requirements for start-up procedure.....	10
201.7. 9.2.9.101* Additional requirements for operating instructions	10
201.7. 9.2.13.101 * Additional requirements for maintenance.....	11
201.7. 9.2.14.101 * Additional requirements for ACCESSORIES, supplementary equipment, used material	11
201.7. 9.2.15.101* Additional requirements for environmental protection	11
201.7. 9.3.101 * Additional requirements for technical description.....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	12
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	12
201.10 Protection against unwanted and excessive radiation HAZARDS	12
201.11 Protection against excessive temperatures and other HAZARDS	12
201.11. 6.4 Leakage	12
201.11. 6.5 * Ingress of water or particulate matter into ME EQUIPMENT or ME SYSTEMS	13
201.11. 6.6 * Cleaning and disinfection of ME EQUIPMENT or ME SYSTEMS	13
201.11. 6.7 Sterilization of ME EQUIPMENT or ME SYSTEMS	13

201.11. 6.8	Compatibility with substances used with ME EQUIPMENT.....	13
201.11. 8.101	Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT.....	14
201.11. 8.101.1	* Supply failure TECHNICAL ALARM CONDITION.....	14
201.11. 8.101.2	* Settings and data storage following short interruptions or automatic switchover	14
201.11. 8.101.3	* Operation following long interruptions.....	14
201.11. 8.101.4	* RESERVE ELECTRICAL POWER SOURCE	14
201.11. 8.101.5	* RESERVE ELECTRICAL POWER SOURCE for transport outside a healthcare facility	15
201.12	Accuracy of controls and instruments and protection against hazardous outputs.....	15
201.12. 1	Accuracy of controls and instruments	15
201.12. 1.101	* Measurement accuracy.....	15
201.12. 1.101.1	General	15
201.12. 1.101.2	* DRIFT of MEASUREMENT ACCURACY	16
201.12. 1.101.3	* MEASUREMENT ACCURACY of GAS READINGS for gas mixtures	17
201.12. 1.102	* TOTAL SYSTEM RESPONSE TIME and rise time	17
201.12. 1.103	* Indication of units of measure for GAS READINGS	18
201.12. 1.104	* Indication of operating mode	19
201.13	HAZARDOUS SITUATIONS and fault conditions	19
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	19
201.15	Construction of ME EQUIPMENT.....	19
201.15. 3.5.101	* Additional requirements for rough handling.....	19
201.15. 3.5.101.1	* Shock and vibration	19
201.15. 3.5.101.2	* Shock and vibration for professional transportation	20
201.15. 101	* Mode of operation	21
201.16	ME SYSTEMS	21
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	21
201.101	* Interfering gas and vapour effects	22
201.102	* Gas leakage.....	22
201.103	* Port connector for DIVERTING RGM.....	22
201.104	* Minimum sampling flowrate	23
201.105	* Contamination of breathing systems	23
201.105. 1	Sampling tube	23
201.105. 2	Exhaust tube	23
202	Electromagnetic compatibility — Requirements and tests	23
202.6.2.1.7	* PATIENT simulation.....	23
202.6.2.1.10	Compliance criteria	23
202.6.2.3.1	* Requirements.....	24
206	Usability	24
206.6.2.2.2	Primary operating functions.....	24
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	24
208.6.1.2	* ALARM CONDITION priority.....	24
208.6.5.1	* General requirements.....	26
208.6.6.2.101	* Additional requirements for adjustable ALARM LIMIT	26
208.6.8.5.101	* Additional requirements for ALARM SIGNAL deactivation states, indication and access	26
209	Requirements for environmentally conscious design.....	26

210	Requirements for the development of physiologic closed-loop controllers	26
211	Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the home healthcare environment.....	27
	Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	28
	201.C.1 Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	28
	201.C.4 ACCOMPANYING DOCUMENTS, general	28
	201.C.5 ACCOMPANYING DOCUMENTS, instructions for use.....	29
	201.C.6 ACCOMPANYING DOCUMENTS, technical description	30
	Annex D (informative) Symbols on marking	31
	Annex AA (informative) Particular guidance and rationale	33
	Annex BB (informative) Environmental aspects.....	43
	Annex CC (informative) Test gas mixtures for calibration	45
	Annex DD (informative) Reference to the essential principles	46
	Bibliography.....	48