

ISO 14117:2012-07 (E)

Active implantable medical devices - Electromagnetic compatibility - EMC test protocols for implantable cardiac pacemakers, implantable cardioverter defibrillators and cardiac resynchronization devices

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions, symbols and abbreviations	1
4 Test requirements for the frequency band 0 Hz – 3 000 MHz	3
4.1 General	3
4.2 Induced lead current	4
4.3 Protection from persisting malfunction attributable to ambient electromagnetic fields	11
4.4 Temporary response to continuous wave sources	21
4.5 Protection from sensing EMI as cardiac signals	24
4.6 Protection from static magnetic fields of flux density up to 1 mT	34
4.7 Protection from static magnetic fields of flux density up to 50 mT	35
4.8 Protection from AC magnetic field exposure in the range of 1 kHz to 140 kHz	36
4.9 Test requirements for the frequency range of 450 MHz – 3 000 MHz	37
5 Testing above frequency of 3 000 MHz	41
6 Protection of devices from EM fields encountered in a therapeutic environment	41
6.1 Protection of the device from damage caused by high-frequency surgical exposure	41
6.2 Protection of the device from damage caused by external defibrillators	42
7 Additional accompanying documentation	46
7.1 Disclosure of permanently programmable sensitivity settings	46
7.2 Descriptions of reversion modes	46
7.3 Known potential hazardous behaviour	46
Annex A (informative) Rationale	47
Annex B (informative) Rationale for test frequency ranges	58
Annex C (informative) Code for describing modes of implantable generators	63
Annex D (normative) Interface circuits	65
Annex E (informative) Selection of capacitor Cx	70
Annex F (normative) Calibration of the injection network (Figure D.5)	71
Annex G (normative) Torso simulator	73
Annex H (normative) Dipole antennas	76
Annex I (normative) Pacemaker/ICD programming settings	78

Annex J (normative) Simulated cardiac signal	80
Annex K (normative) Calculation of net power into dipole antenna	81
Annex L (informative) Loop area calculations	85
Annex M (informative) Correlation between levels of test voltages used in this International Standard and strengths of radiated fields	90
Bibliography	96