

ISO 4037-2:2019 (E)

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV

Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Standard instrument
4.1	General
4.2	Calibration of the standard instrument
4.3	Energy dependence of the response of the standard instrument
5	Conversion from the measured quantity air kerma, K_a , to the required phantom related measuring quantity
5.1	General
5.2	Determination of conversion coefficients
5.2.1	General
5.2.2	Calculation of conversion coefficients from spectral fluence
5.3	Validation of reference fields and of listed conversion coefficients using dosimetry
6	Direct calibration of the reference field in terms of the required phantom related measuring quantity
7	Measurement procedures applicable to ionization chambers
7.1	Geometrical conditions
7.2	Chamber support and stem scatter
7.3	Location and orientation of the standard chamber
7.4	Measurement corrections
7.4.1	General
7.4.2	Corrections for air temperature, pressure and humidity variation from reference calibration conditions
7.4.3	Corrections for radiation-induced leakage, including ambient radiation
7.4.4	Incomplete ion collection
7.4.5	Beam non-uniformity
8	Additional procedures and precautions specific to gamma radiation dosimetry using radionuclide sources
8.1	Use of certified source output
8.2	Use of electron equilibrium caps
8.3	Radioactive source decay
8.4	Radionuclide impurities
8.5	Interpolation between calibration positions
9	Additional procedures and precautions specific to X-radiation dosimetry
9.1	Variation of X-radiation output
9.2	Monitor
9.3	Adjustment of air kerma rate

10 Dosimetry of reference radiation at photon energies between 4 MeV and 9 MeV

- 10.1 Dosimetric quantities**
- 10.2 Measurement of the dosimetric quantities**
 - 10.2.1 General**
 - 10.2.2 Air kerma (rate)**
 - 10.2.3 Phantom related operational quantities $H^*(10)$, $H_p(10)$, $H'(3)$ and $H_p(3)$**
- 10.3 Measurement geometry**
- 10.4 Monitor**
- 10.5 Determination of air kerma (rate) free-in-air**
 - 10.5.1 General**
 - 10.5.2 Measurement conditions**
 - 10.5.2.1 Choice and positioning of detector**
 - 10.5.2.2 Electron equilibrium**
 - 10.5.3 Direct measurement with an ionization chamber**
 - 10.5.4 Determination of air kerma (rate) from photon fluence (rate)**
 - 10.5.4.1 General**
 - 10.5.4.2 Determination of air kerma from photon fluence measurements**

11 Uncertainty of measurement

- 11.1 General**
- 11.2 Components of uncertainty**
 - 11.2.1 General**
 - 11.2.2 Uncertainties in the calibration of a secondary standard**
 - 11.2.3 Uncertainties in the measurements of the reference radiation due to the standard instrument and its use**
 - 11.2.3.1 Uncertainties determined by type A methods**
 - 11.2.3.2 Uncertainties determined by type B methods**
- 11.3 Statement of uncertainty**

Annex A (normative) Technical details of the instruments and their operation

- A.1 General**
- A.2 Operational details of the standard instrument**
 - A.2.1 General**
 - A.2.2 Operation of the standard instrument**
 - A.2.3 Stability check**
 - A.2.4 Warm-up and response times**
 - A.2.5 Zero-setting**
 - A.2.6 Number of readings**
 - A.2.7 Non-constant response value due to instrument scale and range**
 - A.2.8 Shutter transit time**
- A.3 Procedures applicable to ionization chambers**
 - A.3.1 General**
 - A.3.2 Ionization chamber assembly calibrated separately from measuring assembly**
 - A.3.3 Influence of the direction of radiation incidence on the response of the ionization chamber**
 - A.3.4 Measurement of the effect of leakage**
 - A.3.5 Measurement corrections**
 - A.3.5.1 General**
 - A.3.5.2 Zero shift**
 - A.3.5.3 Corrections for electrical leakage**

Annex B (informative) Measurement of photon spectra

- B.1 Experimental set-up for the measurement of pulse height spectra**
- B.2 Unfolding of pulse height spectra to determine the spectral fluence**