

ISO 14708-6:2019 (E)

Implants for surgery — Active implantable medical devices — Part 6: Particular requirements for active implantable medical devices intended to treat tachyarrhythmia (including implantable defibrillators)

Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Symbols and abbreviated terms
5	General requirements for non-implantable parts
5.1	General requirements for non-implantable parts
5.2	General requirements for software
5.3	Usability of non-implantable parts
5.4	Data security and protection from harm caused by unauthorized information tampering
5.5	General requirements for risk management
5.6	Misconnection of parts of the active implantable medical device
6	Measurement of implantable pulse generator and lead characteristics
6.1	Measurement of implantable pulse generator characteristics
6.1.1	General considerations
6.1.2	Measurement of the bradyarrhythmia characteristics
6.1.3	Measurement of ICD output voltage
6.1.4	Measurement of delivered CD pulse energy
6.1.5	Measurement of the antitachyarrhythmia pacing pulse amplitude
6.1.6	Measurement of the sensitivity of an implantable pulse generator with automatic sensitivity control
6.1.7	Charge time
6.1.8	Capacitor formation (capacitor maintenance)
6.2	Measurement of the electrical characteristic of a sensing/pacing lead
7	General arrangement of the packaging
8	General markings for active implantable medical devices
9	Markings on the sales packaging
10	Construction of the sales packaging
11	Markings on the sterile pack
12	Construction of the non-reusable pack
13	Markings on the active implantable medical device
14	Protection from unintended biological effects being caused by the active implantable medical device

- 15 Protection from harm to the patient or user caused by external physical features of the active implantable medical device
- 16 Protection from harm to the patient caused by electricity
- 17 Protection from harm to the patient caused by heat
 - 17.1 Protection from harm to the patient caused by heat
 - 17.2 Active implantable medical device intended to supply heat
- 18 Protection from ionizing radiation released or emitted from the active implantable medical device
- 19 Protection from unintended effects caused by the device
- 20 Protection of the device from damage caused by external defibrillators
- 21 Protection of the device from changes caused by high-power electrical fields applied directly to the patient
- 22 Protection of the active implantable medical device from changes caused by miscellaneous medical treatments
- 23 Protection of the active implantable medical device from mechanical forces
- 24 Protection of the active implantable medical device from damage caused by electrostatic discharge
- 25 Protection of the active implantable medical device from damage caused by atmospheric pressure changes
- 26 Protection of the active implantable medical device from damage caused by temperature changes
- 27 Protection of the active implantable medical device from electromagnetic non-ionizing radiation
- 28 Accompanying documentation
- Annex A (informative) Relationship between the fundamental principles in ISO/TR 14283 and the clauses of this document
- Annex B (informative) Notes on ISO 14708-6
 - B.1 General
 - B.2 Notes on specific clauses and subclauses
- Annex C (informative) Code for describing modes of implantable pulse generators
 - C.1 General
 - C.2 The code
 - C.3 Usage

Page count: 66