ISO 14117:2019-09 (E)

Active implantable medical devices - Electromagnetic compatibility - EMC test protocols for implantable cardiac pacemakers, implant able cardioverter defibrillators and cardiac resynchronization devices

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Terms, definitions, symbols and abbreviated terms 1
3.1 Terms and definitions 1
3.2 Acronyms and abbreviations 3

4 Test requirements for the frequency band 0 Hz 3 000 MHz 4
4.1 General requirements for all devices 4
4.2 Induced lead current 5
4.2.1 General requirements 5
4.2.2 Pacemakers and CRT-P devices 5
4.2.3 ICDs and CRT-D devices 9
4.3 Protection from persisting malfunction attributable to ambient electromagnetic fields 12
4.3.1 General requirements 12
4.3.2 Pacemaker and CRT-P devices 12
4.3.3 ICDs and CRT-D devices 17
4.4 Protection from malfunction caused by temporary exposure to CW sources 23
4.4.1 Pacemaker and CRT-P device' response to temporary continuous wave sources in the frequency range 16,6 Hz to 167 kHz 23
4.4.2 ICDs and CRT-D devices 25
4.5 Protection from sensing EMI as cardiac signals 26
4.5.1 General requirements 26
4.5.2 Protection from sensing EMI as cardiac signals in the frequency range of 16,6 Hz to 150 kHz 27
4.5.3 Protection from sensing EMI as cardiac signals in the frequency range of 150 kHz to 10 MHz 30
4.5.4 Protection from sensing EMI as cardiac signals in the frequency range of 10 MHz to 385 MHz 33
4.6 Protection from static magnetic fields of flux density up to 1 mT 35
4.6.1 General requirements 35
4.6.2 Pacemakers and CRT-P devices 35
4.6.3 ICDs and CRT-D devices 36
4.7 Protection from static magnetic fields of flux density up to 50 mT 37
4.7.1 General requirements 37
4.7.2 Pacemakers and CRT-P devices 37
4.7.3 ICDs and CRT-D devices 37
4.8 Protection from AC magnetic field exposure in the range of 1 kHz to 140 kHz 37
4.8.1 General requirements 37
4.8.2 Pacemakers and CRT-P devices 37
4.8.3 ICDs and CRT-D devices 38
4.9 Test requirements for the frequency range of 385 MHz 3 000 MHz 38
4.9.1 General requirements 38
4.9.2 Test setup 39

4.9.3 Test procedure 40
4.9.4 Performance criteria 42
4.10 Transient exposure to stationary low-frequency electromagnetic field sources in the frequency range 16,6 Hz to 167 kHz 43

5 Testing above frequency of 3 000 MHz 43

6 Protection of devices from EM fields encountered in a therapeutic environment 43
6.1 Protection of the device from damage caused by high-frequency surgical exposure 43
6.1.1 General requirements 43

6.1.2 Pacemakers and CRT-P devices 44
6.1.3 ICDs and CRT-D devices 44
6.2 Protection of the device from damage caused by external defibrillators 45
6.2.1 General requirements 45
6.2.2 Pacemakers and CRT-P devices 45
6.2.3 ICDs and CRT-D devices 48

7 Additional accompanying documentation 49
7.1 Disclosure of permanently programmable sensitivity settings 49
7.2 Descriptions of reversion modes 49
7.3 Known potential hazardous behaviour 49
7.4 Minimum separation distance from hand-held transmitters 49

Annex A (informative) Rationale 50

Annex B (informative) Rationale for test frequency ranges 63

Annex C (informative) Code for describing modes of implantable generators 64

Annex D (normative) Interface circuits 66

Annex E (informative) Selection of capacitor Cx 71

Annex F (normative) Calibration of the injection network (Figure D.5) 74

Annex G (normative) Torso simulator 76

Annex H (normative) Dipole antennas 80

Annex I (normative) Pacemaker/ICD programming settings 82

Annex J (normative) Simulated cardiac signal 84

Annex K (normative) Calculation of net power into dipole antenna 85

Annex L (informative) Loop area calculations 90

Annex M (informative) Correlation between levels of test voltages used in this document and strengths of radiated fields 96

Annex N (informative) Connections to DUTs having ports with more than two electrode connections 104

Annex O (informative) Example method for evaluation of transient and permanent malfunction of a CIED due to temporary exposure to low frequency (<167 kHz) electromagnetic fields 127

Bibliography 132