ISO 27186:2026-07 (E)

Active implantable medical devices - Four-pole connector system for implantable cardiac rhythm management devices - Dimensional and test requirements

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 2

3 Termsanddefinitions 2

4 Requirements 5
4.1 General 5
4.2 Lead connector physical requirements 5
4.2.1 Dimensions 5
4.2.2 Materials 10
4.2.3 Lead connector electrical connections 10
4.2.4 Lead marking 10
4.2.5 Lead package labels and literature 12
4.3 Lead connector functional requirements 12
4.3.1 Functional fit check 12
4.3.2 Tensile loads 13
4.3.3 Deformation due to pin contact forces 13
4.3.4 Deformation due to ring contact forces 13
4.3.5 Seal zone requirement 14
4.3.6 Electrical isolation requirement 15
4.3.7 Dielectric strength requirement 15
4.3.8 Current-carrying requirement 15
4.3.9 Corrosion and environmental requirement 15
4.3.10 Lead connector insertion force 15
4.3.11 Lead connector fatigue strength 15
4.4 Connector cavity physical requirements 15
4.4.1 Dimensions 15
4.4.2 Connector cavity electrical connections 17
4.4.3 Connector cavity and pulse generator marking 18
4.4.4 Pulse generator labels and literature 18
4.5 Connector cavity functional requirements 19
4.5.1 Insertion force 19
4.5.2 Retention force 20
4.5.3 Withdrawal force 20
4.5.4 Ring contact load 20
4.5.5 Seal zone load requirement 21
4.5.6 Electrical isolation requirement 21
4.5.7 Dielectric strength requirement 21
4.5.8 Current-carrying requirement (high-voltage connector cavity) 21
4.5.9 Contact resistance and stability 21

Annex A (normative) Electrical isolation test method 22

Annex B (informative) Rationale for electrical isolation test 27

Annex C (normative) Dielectric strength test method 29

Annex D (informative) Rationale for dielectric strength test 34

Annex E (normative) Current-carrying test methods 38

Annex F (informative) Rationale for current-carrying tests 44

Annex G (informative) Guidance for evaluating high cycle fatigue of four-pole connectors 46

Annex H (informative) Guidance on selecting materials for lead connector seal zones 47

Annex I (informative) Seal zone creep 49

ISO 27186:2026(en) Annex J (informative) Contact resistance stability -- Example test method 53

Annex K (informative) Rationale for contact resistance stability example test 56

Annex L (informative) Guidance on selecting materials for electrical contacts 58

Annex M (normative) Lead connector contact material requirements and test methods 60

Annex N (informative) Rationale for lead connector contact material requirements 63

Annex O (informative) Rationale for standardization and standard requirements 69

Annex P (informative) Connector-related products and adaptors 75

Annex Q (normative) Lead connector insertion force test method 77

Annex R (informative) Rationale for lead connector insertion force test 86

Annex S (normative) Connector cavity insertion and withdrawal test method 93

Annex T (informative) Rationale for connector cavity insertion and withdrawal test 99

Bibliography 107