

DIN EN ISO 389-5:2007-05 (E)

**Acoustics - Reference zero for the calibration of audiometric equipment - Part 5:
Reference equivalent threshold sound pressure levels for pure tones in the
frequency range 8 kHz to 16 kHz (ISO 389-5:2006)**

Contents	Page
Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Specifications	6
Annex A (informative) Notes on the derivation of the reference equivalent threshold sound pressure levels for audiometric earphones in the frequency range from 8 kHz to 16 kHz	7
Annex B (informative) RETSPL values for Koss HV/1A	8
Bibliography	10