

# ISO 389-5:2006-11 (E)

Acoustics - Reference zero for the calibration of audiometric equipment - Part 5:  
Reference equivalent threshold sound pressure levels for pure tones in the  
frequency range 8 kHz to 16 kHz

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
1 <b>Scope</b> .....		1
2 <b>Normative references</b> .....		1
3 <b>Terms and definitions</b> .....		1
4 <b>Specifications</b> .....		2
<b>Annex A (informative) Notes on the derivation of the reference equivalent threshold sound pressure levels for audiometric earphones in the frequency range from 8 kHz to 16 kHz</b> ....		3
<b>Annex B (informative) RETSPL values for Koss HV/1A</b> .....		4
<b>Bibliography</b> .....		6