ISO 389-5:2006-11 (E)

Acoustics - Reference zero for the calibration of audiometric equipment - Part 5: Reference e quivalent threshold sound pressure levels for pure tones in the frequency range 8 kHz to 16 kHz

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 Specifications 2

Annex A (informative) Notes on the derivation of the reference equivalent threshold sound pressure levels for audiometric earphones in the frequency range from 8 kHz to 16 kHz 3

Annex B (informative) RETSPL values for Koss HV/1A 4

Bibliography 6