

DIN EN 14624:2012-04 (E)

Performance of portable leak detectors and of room monitors for halogenated refrigerants

Contents		Page
Foreword		4
1	Scope	5
2	Normative references	5
3	Terms and definitions	5
4	Symbols and abbreviations	7
5	Types of portable leak detectors and room monitors	7
6	General requirements for all portable leak detectors and room monitors	7
6.1	Gas type	7
6.2	Capabilities	7
6.3	Output signal	8
6.4	Response time	8
6.5	Recovery time (clean up time)	8
6.6	Repeatability	8
6.7	Calibration	8
7	Specific requirements for locating portable leak detectors	8
7.1	General	8
7.2	Lowest detectable leakage rate threshold when the leak detector is stationary	9
7.3	Lowest detectable leakage rate threshold value when the detector probe is moving	9
7.4	Leak detection in a contaminated environment	9
8	Specific requirements for room monitors	10
8.1	General	10
8.2	Lowest detectable concentration (lower detection limit)	10
8.3	Highest tolerable concentration (upper detection limit)	10
8.4	Upper and lower alarm thresholds	10
8.5	Response time for alarm trigger	10
9	Calibration leaks	10
9.1	General	10
9.2	Calibration gases	10
10	Test apparatus	11
10.1	General	11
10.2	Apparatus no. 1: detector probe stationary at the orifice of a calibration leak	11
10.3	Apparatus no. 2: detector probe moved with defined speed and distance in front of a calibration leak	11
10.4	Apparatus no. 3: Chamber with monitored concentration	12
11	Performance tests	12
11.1	Performance test of locating leak detector	12
11.1.1	Test no. 1: Static detection	12
11.1.2	Test no. 2: dynamic detection limit	13
11.1.3	Test no. 3: detection limit in contaminated environment	13
11.1.4	Test no. 4: recovery time	13

11.2	Performance tests of room monitor	13
11.2.1	General	13
11.2.2	Test a) to determine the upper and lower detection limits	13
11.2.3	Test b) to determine response time	14
11.2.4	Test c) to determine recovery time from exposure to a large concentration	14
12	Instrument characteristics and reporting of test results	15
12.1	Characteristics of locating leak detector	15
12.2	Characteristics of room monitors	15
13	Technical data	16
13.1	Locating leak detectors	16
13.2	Measuring leak detectors	16
13.3	All portable leak detectors	16
13.4	Room monitor	16
Annex A (informative) Conversion factors of leakage rate units		18
Annex B (informative) Correlation between test gas concentration and leakage rate		19
Annex C (informative) Guidelines for monitoring a machinery room or space for gas leaks		23