

# ISO 648:2008-10 (E)

## Laboratory glassware - Single-volume pipettes

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
<b>1</b>	<b>Scope .....</b>	<b>1</b>
<b>2</b>	<b>Normative references .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Basis of adjustment .....</b>	<b>2</b>
<b>4.1</b>	<b>Unit of volume .....</b>	<b>2</b>
<b>4.2</b>	<b>Delivery volume .....</b>	<b>2</b>
<b>4.3</b>	<b>Reference temperature .....</b>	<b>2</b>
<b>5</b>	<b>Types and classes of accuracy .....</b>	<b>2</b>
<b>5.1</b>	<b>Classes of accuracy .....</b>	<b>2</b>
<b>5.2</b>	<b>Types of single-volume pipette .....</b>	<b>2</b>
<b>6</b>	<b>Maximum permissible errors .....</b>	<b>3</b>
<b>7</b>	<b>Construction .....</b>	<b>3</b>
<b>7.1</b>	<b>Material .....</b>	<b>3</b>
<b>7.2</b>	<b>Shape .....</b>	<b>3</b>
<b>7.3</b>	<b>Bulb .....</b>	<b>3</b>
<b>7.4</b>	<b>Dimensions .....</b>	<b>4</b>
<b>7.5</b>	<b>Top of pipette .....</b>	<b>5</b>
<b>7.6</b>	<b>Delivery jet .....</b>	<b>5</b>
<b>7.7</b>	<b>Delivery time .....</b>	<b>5</b>
<b>7.8</b>	<b>Waiting time .....</b>	<b>6</b>
<b>8</b>	<b>Graduation line(s) .....</b>	<b>6</b>
<b>9</b>	<b>Setting of the meniscus .....</b>	<b>6</b>
<b>10</b>	<b>Marking .....</b>	<b>6</b>
<b>11</b>	<b>Visibility of graduation lines, figures and inscriptions .....</b>	<b>7</b>
<b>12</b>	<b>Colour coding .....</b>	<b>7</b>
Bibliography .....		8