DIN EN 14212:2025-12 (E)

Ambient air - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence

Contents

Page

European foreword ... 8

1 Scope ... 9

2 Normative references ... 10

3 Terms and definitions ... 11

4 Abbreviated terms ... 16

5 Principle ... 17
5.1 General ... 17
5.2 Measuring principle ... 17
5.3 Type testing ... 18
5.4 Field operation and quality control ... 18

6 Sampling ... 18
6.1 General ... 18
6.2 Sampling location ... 19
6.3 Sampling system ... 19
6.4 Control and regulation of sample flow rate ... 20
6.5 Sampling pump for the manifold ... 20

7 Analyser equipment ... 21
7.1 General ... 21
7.2 Selective traps for interfering agents ... 21
7.3 Optical assembly ... 21
7.4 Pressure measurement ... 21
7.5 Flow rate indicator ... 22
7.6 Sampling pump for the analyser ... 22
7.7 Internal sulfur dioxide span source ... 22
7.8 Particle filter ... 22

8 Type testing of ultraviolet fluorescence sulfur dioxide analysers ... 22
8.1 General ... 22
8.2 Relevant performance characteristics and criteria ... 23

Table 1 — Relevant performance characteristics and criteria ... 24
8.3 Design change ... 26
8.4 Procedures for determination of the performance characteristics during the laboratory test ... 26

Table 2 — Set points and stability of test parameters ... 27

Table 3 — Methods for preparation of test gases ... 27

Table 4 — Specification for purity of test gas ... 28

Table 5 — Specification for purity of zero gas for interferents testing [9] ... 28

Table 6 — Specification for purity of zero gas for other tests [9] ... 28

Figure 1 — Diagram illustrating the response time ... 30

Figure 2 — Concentration variations for the averaging effect test ... 37
8.5 Determination of the performance characteristics during the field test ... 38
8.6 Type testing and uncertainty calculation ... 42

9 Field operation and ongoing quality control ... 42

9.1 General 42
9.2 Suitability evaluation 43
Table 7 — Site-specific conditions to be evaluated 43
9.3 Initial installation 44
9.4 Ongoing quality assurance/quality control 46
Table 8 — Required frequency of calibration, checks and maintenance 47
9.5 Calibration of the analyser 49
Figure 3 — Flow scheme for performance of evaluation of effects of violation of performance requirements and possibilities for data correction 52
9.6 Checks 53
9.7 Maintenance 56
9.8 Data handling and data reports 57
9.9 Measurement uncertainty 57
10 Expression of results 58
11 Test reports and documentation 58
11.1 Type testing 58
11.2 Field operation 59
Annex A (normative) Test of lack of fit 61
A.1 Establishment of the regression line 61
A.2 Calculation of the residuals of the averages 61
Annex B (informative) Sampling equipment 63
Figure B.1 — Sampling layout with a main sampling manifold 63
Figure B.2 — Sampling layout with individual lines 64
Annex C (informative) Ultraviolet fluorescence analyser 65
Figure C.1 — Schematic diagram of a UV fluorescence sulfur dioxide analyser 65
Annex D (informative) Manifold testing 66
D.1 Procedure for applying test gas 66
Figure D.1 — Schematic diagram of manifold testing equipment 67
D.2 Procedure for the cross test 67
D.2.1 General 67
D.2.2 Initial stage 67
D.2.3 Stage 1 68
D.2.4 Stage 2 68
D.2.5 Data Processing 68
D.2.6 Evaluation 68
Annex E (normative) Type testing 70
E.1 Type testing and uncertainty calculation 70
E.1.1 Type testing 70
E.1.2 Uncertainty calculation 70
E.2 Type testing Requirement a) 70

Table E.1 — Relevant performance characteristics and criteria 71
E.3 Type testing Requirement b) 72
E.3.1 General 72
Table E.2 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory tests 73
Table E.3 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory tests 73
E.3.2 Calculation of standard uncertainties 75
E.3.2.1 General 75
E.3.2.2 Repeatability at zero 75
E.3.2.3 Repeatability at the hourly limit value of sulfur dioxide 75
E.3.2.4 Lack of fit 76
E.3.2.5 Influence quantities 76
E.3.2.5.1 General 76
E.3.2.5.2 Sample gas pressure 77
E.3.2.5.3 Sample gas temperature 77
E.3.2.5.4 Surrounding temperature 78
E.3.2.5.5 Electrical voltage 78
E.3.2.5.6 Water vapour 79
E.3.2.5.7 Other interferents 79
E.3.2.5.8 Averaging effect 80
E.3.2.5.9 Calibration gas 81
E.3.2.5.10 Difference sample/calibration port 81
E.3.3 Example calculation 82
Table E.4 — Example calculation 82
E.4 Type testing Requirement c) 84
Table E.5 — Relevant performance characteristics and criteria 84
E.5 Type testing Requirement d) 84
E.5.1 General 84
Table E.6 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests 85
Table E.7 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests 85
E.5.2 Combined standard uncertainty 86
E.5.3 Absolute expanded uncertainty 86
E.5.4 Relative expanded uncertainty 87
E.5.5 Calculation of standard uncertainties 87
E.5.6 Example calculation 89

Table E.8 — Example calculation **89**
Annex F (informative) Calculation of uncertainty in field operation at the hourly limit value **91**
F.1 General **91**
F.2 Combined standard uncertainty **91**
F.3 Standard uncertainties **92**
F.3.1 General **92**
F.3.1.1 General **92**
F.3.1.2 Repeatability at zero **92**
F.3.1.3 Repeatability at the hourly limit value **92**
F.3.1.4 Lack of fit **92**
F.3.2 Influence quantities **93**
F.3.2.1 General **93**
F.3.2.2 Sample gas pressure **93**
F.3.2.3 Sample gas temperature **93**
F.3.2.4 Surrounding temperature **94**
F.3.2.5 Electrical voltage **95**
F.3.3 Interferents **95**
F.3.3.1 General **95**
F.3.3.2 Water vapour **95**
F.3.3.3 Other interferents **96**
F.3.4 Averaging effect **96**
F.3.5 Reproducibility under field conditions **96**
F.3.6 Long-term drift at zero **97**
F.3.7 Long-term drift at level of the hourly limit value **97**
F.3.8 Zero gas **97**
F.3.9 Calibration gas **97**
F.3.10 Difference sample/calibration port **97**
F.4 Example calculation **98**
Table F.1 — Example calculation **98**
Annex G (informative) Calculation of uncertainty in field operation at the daily limit value **100**
G.1 General **100**
G.2 Combined standard uncertainty **100**
G.3 Standard uncertainties **101**
G.3.1 General **101**
G.3.1.1 General **101**
G.3.1.2 Repeatability at zero **101**
G.3.1.3 Repeatability at the daily limit value **102**

G.3.1.4 Lack of fit 102
G.3.2 Influence quantities 102
G.3.2.1 General 102
G.3.2.2 Sample gas pressure 103
G.3.2.3 Sample gas temperature 103
G.3.2.4 Surrounding temperature 104
G.3.2.5 Electrical voltage 104
G.3.3 Interferents 105
G.3.3.1 General 105
G.3.3.2 Water vapour 105
G.3.3.3 Other interferents 106
G.3.4 Averaging effect 107
G.3.5 Zero gas 107
G.3.6 Calibration gas 107
G.3.7 Difference sample/calibration port 107
G.3.8 Reproducibility under field conditions 108
G.3.9 Long-term drift at zero 108
G.3.10 Long-term drift at level of the daily limit value 108
G.4 Example calculation 109
Table G.1 — Example calculation 109
Annex H (informative) Calculation of uncertainty in field operation at the annual critical level 111
H.1 General 111
H.2 Combined standard uncertainty 111
H.3 H.3 Standard uncertainties 113
H.3.1 General 113
H.3.2 Repeatability at zero 113
H.3.3 Repeatability at the annual limit value 113
H.3.4 Lack of fit 114
H.3.5 Influence quantities 114
H.3.5.1 General 114
H.3.5.2 Sample gas pressure 114
H.3.5.3 Sample gas temperature 115
H.3.5.4 Surrounding temperature 115
H.3.5.5 Electrical voltage 116
H.3.5.6 Interferents 117
H.3.5.7 Water vapour 117

H.3.5.8 Other interferents 117
H.3.6 Averaging effect 118
H.3.7 Zero gas 119
H.3.8 Calibration gas 119
H.3.9 Difference sample/calibration port 119
H.3.10 Reproducibility under field conditions 120
H.3.11 Long term drift at zero 120
H.3.12 Long term drift at level of the annual critical level 121
Annex I (informative) Test stand for the test point "sensitivity coefficient of sample gas pressure" 122
Figure I.1 — Test stand for the test point "sensitivity coefficient of sample gas pressure" 122
Annex J (informative) Significant technical changes 123
Table J.1 — Details of significant technical changes between this document and the previous edition 123
Bibliography 124