ISO 8932-2:2026-05 (E)

Meteorology - Radiosonde - Part 2: Laboratory test method for errors in radiosonde humidity sensor calibration

Contents

Page

Foreword ..... **v**

Introduction ..... **vi**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **2**

4 Symbols and subscripts **4**
4.1 Symbols 4
4.2 Subscript 4

5 Technical requirements for the laboratory setup **5**
5.1 General 5
5.1.1 Temperature 5
5.1.2 Pressure 5
5.2 Precision hygrometer 6
5.2.1 Type 6
5.2.2 Installation 6
5.2.3 Operation 7
5.3 Humidity generator 8
5.3.1 Type 8
5.3.2 Installation 8
5.3.3 Operation 9

6 Test preparation **10**
6.1 Laboratory environmental conditions 10
6.2 Preparation of the radiosonde 10
6.3 Examination of the laboratory setup 11
6.3.1 General 11
6.3.2 Examination of the dry gas generator 11
6.3.3 Examination of the liquid bath and climate chamber 11
6.3.4 Examination of the measurement system for the calculation of the reference relative humidity in the test cell 11
6.3.5 Examination of the radiosonde measurement software and the control software for the reference relative humidity 11
6.4 Installation of the radiosonde 11

7 Test methods and procedures **11**
7.1 Operation 11
7.1.1 Purging the test cell 11
7.1.2 Temperature control of the test cell 12
7.1.3 Humidity from the humidity generator 12
7.1.4 Calculation of the reference relative humidity using the humidity generator 13
7.1.5 Calculation of the reference relative humidity using the precision hygrometer 13
7.2 Test procedure 13

8 Data processing **14**
8.1 Calculation of the average values 14
8.2 Calculation of the measurement error 14

9 Evaluation of measurement uncertainty **14**
9.1 General 14

9.2 Uncertainty evaluation for the reference relative humidity, $u(h_{ref})$ 15
9.2.1 Uncertainty of the reference relative humidity 15
9.3 Uncertainty of the radiosonde relative humidity, $u(h_{rad})$ 17
9.3.1 Uncertainty of the resolution of the radiosonde relative humidity, $u(h_{rad_res})$ 17
9.3.2 Uncertainty of the repeatability of the radiosonde relative humidity, $u(h_{rad_rep})$ 17
9.4 Calculation of the combined standard uncertainty of the measurement error, $u(h_{err})$ 17
9.5 Calculation of expanded uncertainty 18
10 Method for reporting test results **18**
Annex A (informative) **Calculation of the reference relative humidity** **19**
Bibliography **23**