

ISO 23496:2019 (E)

Determination of pH value — Reference buffer solutions for the calibration of pH measuring equipment

Contents

	Foreword
1	Scope
2	Normative references
3	Terms and definitions
4	Basic information
4.1	General
4.2	pH values of primary reference buffer solutions
4.3	pH values of secondary reference buffer solutions
5	Preparation of primary and secondary reference buffer solutions
5.1	General
5.2	Reference buffer solution A, pH(R) = 1,6791 1 The last digit is not significant and therefore printed as subscript.(25 °C), Potassium tetraoxalate dihydrate
5.3	Reference buffer solution B, pH(R) = 3,557 (25 °C), Potassium hydrogen tartrate
5.4	Reference buffer solution C, pH(R) = 4,005 (25 °C), Potassium hydrogen phthalate
5.5	Reference buffer solution D, pH(R) = 6,865 (25 °C), Phosphate
5.6	Reference buffer solution E, pH(R) = 7,413 (25 °C), Phosphate
5.7	Reference buffer solution F, pH(R) = 9,180 (25 °C), Borax
5.8	Reference buffer solution G, pH(R) = 12,454 (25 °C), Calcium hydroxide
5.9	Reference buffer solution H, pH(R) = 3,776 (25 °C), Potassium dihydrogen citrate
5.10	Reference buffer solution I, pH(R) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate
6	Storage and shelf life
7	Examples of pH(R) values of reference buffer solutions
8	Additional properties of reference buffer solutions
Annex A	(informative) Examples of pH(R) values as a function of temperature at temperatures from 55 °C to 95 °C
Annex B	(informative) Volumetric method
B.1	General
B.2	Reference buffer solution A, pH(R) = 1,679 (25 °C), Potassium tetraoxalate dihydrate
B.3	Reference buffer solution B, pH(R) = 3,557 (25 °C), Potassium hydrogen tartrate
B.4	Reference buffer solution C, pH(R) = 4,005 (25 °C), Potassium hydrogen phthalate
B.5	Reference buffer solution D, pH(R) = 6,865 (25 °C), Phosphate
B.6	Reference buffer solution E, pH(R) = 7,413 (25 °C), Phosphate
B.7	Reference buffer solution F, pH(R) = 9,180 (25 °C), Borax
B.8	Reference buffer solution G, pH(R) = 12,454 (25 °C), Calcium hydroxide
B.9	Reference buffer solution H, pH(R) = 3,776 (25 °C), Potassium dihydrogen citrate
B.10	Reference buffer solution I, pH(R) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate