

ISO 21501-1:2009-06 (E)

Determination of particle size distribution - Single particle light interaction methods - Part 1: Light scattering aerosol spectrometer

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Terms and definitions	1
3	Requirements	3
3.1	Size range	3
3.2	Counting efficiency	3
3.2.1	General	3
3.2.2	Lower size limit	4
3.2.3	Upper size limit	4
3.3	Size resolution	4
3.4	Sizing accuracy	5
3.5	Sampling flow rate	5
3.6	Effective detection flow rate	5
3.7	Maximum particle number concentration	5
4	Test method	5
4.1	Size calibration	5
4.2	Effective detection flow rate	6
4.3	Maximum particle number concentration	7
4.4	Size resolution	8
4.5	Counting efficiency	9
Annex A (informative)	Principle of the instruments	11
Annex B (informative)	Particle size standards	18
Annex C (informative)	Effects of the LSAS parameters on the particle size and particle number concentration determination	21
Annex D (informative)	Representative sampling	22
Annex E (informative)	Example of an LSAS calibration with DEMS-classified PSL particles	24
Bibliography		26