

ISO 21501-4:2018 (E)

Determination of particle size distribution — Single particle light interaction methods — Part 4: Light scattering airborne particle counter for clean spaces

Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Principle
5	Basic configuration
6	Requirements
6.1	Size setting error
6.2	Counting efficiency
6.3	Size resolution
6.4	False count
6.5	Maximum particle number concentration
6.6	Sampling flow rate error
6.7	Sampling time error
6.8	Response rate
6.9	Calibration interval
6.10	Reporting of test and calibration results
7	Test and calibration procedures
7.1	Size setting
7.1.1	Evaluation of size setting error
7.1.2	Procedure of size setting
7.2	Evaluation of counting efficiency
7.3	Evaluation of size resolution
7.4	Evaluation of false count
7.5	Estimation of coincidence loss at the maximum particle number concentration
7.6	Evaluation of sampling flow rate error
7.7	Evaluation of sampling time error
7.8	Evaluation of response rate
Annex A	(informative) Counting efficiency
Annex B	(informative) Size resolution
Annex C	(informative) False count
Annex D	(informative) Response rate
Annex E	(informative) Procedure for evaluating the uncertainties of the results of the performance tests
E.1	Basics on measurement uncertainty
E.2	Uncertainty of counting efficiency, η
E.3	Uncertainty of the size setting error, ε
E.4	Consideration of test uncertainties in the requirements for LSAPCs

