

DIN EN ISO 7488:2018-08 (E)

Dentistry - Mixing machines for dental amalgam (ISO 7488:2018)

Contents		Page
European foreword		4
Foreword		5
Introduction		6
1	Scope	7
2	Normative references	7
3	Terms and definitions	7
4	Requirements	8
4.1	Safety	8
4.1.1	Electrical	8
4.1.2	Mechanical	8
4.2	Stability	8
4.3	Sound pressure	9
4.4	Frequency	9
4.4.1	General	9
4.4.2	Variable-power machines	9
4.5	Amplitude	9
4.6	Mixing time	9
4.6.1	Timing device	9
4.6.2	Timing settings	9
4.7	Coherence time	10
4.8	Long-term test	10
5	Sampling	10
5.1	Mixing machine	10
5.2	Mixing-capsules	10
5.2.1	General	10
5.2.2	Mixing-capsule selection	10
5.3	Test components	11
6	Measurement and test methods	11
6.1	Test conditions	11
6.1.1	General	11
6.1.2	Reference charge mass, m_{ref}	11
6.1.3	Minimum charge mass, m_{min}	11
6.1.4	Maximum charge mass, m_{max}	11
6.1.5	Maximum running duration, t_{max}	11
6.1.6	Power supply conditions and apparatus	11
6.1.7	Equipment setup	12
6.2	Visual inspection	12
6.3	Sound pressure	12
6.3.1	Apparatus and setup	12
6.3.2	Procedure	13
6.3.3	Variable-power machines	13

6.4	Frequency tests.....	13
6.4.1	Measurement of mixing-capsule oscillation frequency.....	13
6.5	Working length.....	14
6.5.1	Principle.....	14
6.5.2	Apparatus.....	14
6.5.3	Procedure.....	14
6.6	Amplitude.....	14
6.6.1	Apparatus.....	14
6.6.2	Procedure.....	14
6.7	Mixing time.....	15
6.7.1	Apparatus.....	15
6.7.2	Procedure.....	15
6.8	Coherence time.....	15
6.8.1	Apparatus and materials.....	15
6.8.2	Procedure.....	15
6.9	Long-term test.....	16
7	Instructions for use.....	17
7.1	General.....	17
7.2	Identification.....	17
7.3	Safety.....	17
7.4	User maintenance.....	17
7.5	Other maintenance.....	17
7.6	Mixing-capsule requirements.....	17
7.7	Mixing-capsule amplitude.....	17
7.8	Operational limitations.....	18
7.8.1	Mixing machine limitations.....	18
7.8.2	Mass and frequency.....	18
8	Packaging.....	18
9	Marking.....	18
9.1	On the mixing-machine.....	18
9.2	On the packaging.....	18
	Bibliography.....	19