

# ISO 9680:2026-05 (E)

## Dentistry - Operating lights

---

<b>Contents</b>		<b>Page</b>
<b>Foreword</b>		<b>v</b>
<b>Introduction</b>		<b>vi</b>
<b>1</b>	<b>Scope</b>	<b>1</b>
<b>2</b>	<b>Normative references</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions</b>	<b>2</b>
<b>4</b>	<b>Requirements</b>	<b>2</b>
4.1	Optical requirements	2
4.1.1	General	2
4.1.2	Adjustable level of illuminance	2
4.1.3	Illumination pattern	2
4.1.4	Illuminance in patient's eyes	4
4.1.5	Chromatic uniformity	4
4.1.6	Chromaticity and correlated colour temperature	4
4.1.7	Shadow	5
4.1.8	Colour fidelity	5
4.1.9	Actinic UV hazard exposure for the skin and eye	6
4.1.10	Near-UV hazard exposure for the eye	6
4.1.11	Retinal blue light hazard exposure	6
4.1.12	Single fault condition for photobiological safety	6
4.1.13	Irradiance (300 nm to 2 500 nm)	6
4.1.14	Compatibility with light-activated restorative materials	7
4.2	Mechanical requirements	7
4.2.1	Operating controls	7
4.2.2	Rotational positioning	7
4.2.3	Stability after positioning	7
4.2.4	Expelled parts	7
4.3	Processing	7
4.4	Electrical requirements	8
4.5	Usability	8
<b>5</b>	<b>Sampling</b>	<b>8</b>
<b>6</b>	<b>Testing</b>	<b>8</b>
6.1	General	8
6.2	Visual inspection	8
6.3	Optical tests	8
6.3.1	Test set-up	8
6.3.2	Level of illuminance and illuminance pattern	9
6.3.3	Illuminance uniformity	9
6.3.4	Illuminance in the patient's eyes	9
6.3.5	Chromatic uniformity	9
6.3.6	Chromaticity and correlated colour temperature	10
6.3.7	Shadow	10
6.3.8	Colour fidelity	11
6.3.9	Actinic UV hazard exposure for the skin and eye	11
6.3.10	Near-UV hazard exposure for the eye	12
6.3.11	Retinal blue light hazard exposure	12
6.3.12	Irradiance (300 nm to 2 500 nm)	14
6.3.13	Compatibility with light-activated restorative materials	14

<b>7</b>	<b>Manufacturer's instructions</b> .....	<b>14</b>
7.1	Documents.....	14
7.2	Instructions for use.....	14
7.3	Technical description.....	15
<b>8</b>	<b>Packaging</b> .....	<b>15</b>
8.1	General.....	15
8.2	Identification.....	15
<b>9</b>	<b>Marking</b> .....	<b>15</b>
9.1	Marking on the outside of operating lights.....	15
9.2	Graphical symbols.....	16
<b>Annex A</b>	<b>(informative) Additional information for retinal blue light hazard exposure test method</b> .....	<b>17</b>
<b>Annex B</b>	<b>(normative) Normalized absorbance of camphorquinone</b> .....	<b>22</b>
<b>Bibliography</b>	.....	<b>25</b>