

# DIN EN ISO 81060-1:2012-08 (E)

## Non-invasive sphygmomanometers - Part 1: Requirements and test methods for non-automated measurement type (ISO 81060-1:2007)

---

<b>Contents</b>		<b>Page</b>
Foreword .....		4
Introduction .....		5
1 * Scope .....		6
2 Normative references .....		6
3 Terms and definitions .....		7
4 Identification and marking .....		10
4.1 * Units of measurement .....		10
4.2 * Legibility of markings .....		10
4.3 * Durability of markings .....		10
4.4 * Marking of non-automated sphygmomanometer .....		10
4.5 * Usability of reading .....		11
4.6 Marking of the cuff .....		11
4.7 Marking of the non-automated sphygmomanometer packaging .....		11
5 General requirements for testing non-automated sphygmomanometers .....		12
5.1 * Type tests .....		12
5.2 * Representative sample .....		12
5.3 Environmental conditions .....		12
5.4 Repairs and modifications .....		12
5.5 * Humidity preconditioning treatment .....		12
6 General requirements .....		13
6.1 General .....		13
6.2 Electrical safety .....		13
6.3 Mechanical safety .....		13
6.4 Mechanical strength .....		13
7 Requirements .....		15
7.1 Pressure indicating means .....		15
7.2 Pneumatic system .....		16
7.3 * Tamper proofing or unauthorized access .....		19
7.4 Dynamic response in normal use .....		19
8 Additional requirements for non-automated sphygmomanometer with mercury manometer .....		20
8.1 * Internal diameter of the tube containing mercury .....		20
8.2 * Portable non-automated sphygmomanometer .....		20
8.3 * Prevention of mercury spillage during transport .....		20
8.4 * Prevention of mercury spillage in normal use .....		20
8.5 Quality of the mercury .....		21
9 Non-automated sphygmomanometers with aneroid manometer .....		21
9.1 * Scale mark at zero .....		21
9.2 * Zero .....		21
9.3 Hysteresis error .....		21
9.4 * Construction and materials .....		22

<b>10</b>	<b>Cleaning, sterilization and disinfection .....</b>	<b>22</b>
<b>10.1</b>	<b>Reusable non-automated sphygmomanometer and parts .....</b>	<b>22</b>
<b>10.2</b>	<b>Non-automated sphygmomanometer and parts requiring processing before use .....</b>	<b>22</b>
<b>10.3</b>	<b>Non-automated sphygmomanometer and parts delivered sterile .....</b>	<b>23</b>
<b>11</b>	<b>Biocompatibility .....</b>	<b>23</b>
<b>12</b>	<b>Information supplied by the manufacturer .....</b>	<b>23</b>
<b>12.1</b>	<b>Accompanying document .....</b>	<b>23</b>
<b>12.2</b>	<b>Instructions for use .....</b>	<b>23</b>
<b>12.3</b>	<b>Technical description .....</b>	<b>26</b>
	<b>Annex A (informative) Rationale and guidance .....</b>	<b>28</b>
	<b>Annex B (informative) Advice regarding non-automated sphygmomanometers with a mercury manometer .....</b>	<b>36</b>
	<b>Annex C (informative) Environmental aspects .....</b>	<b>37</b>
	<b>Annex D (informative) Reference to the essential principals .....</b>	<b>38</b>
	<b>Annex E (informative) Terminology -- Alphabetized index of defined terms .....</b>	<b>40</b>
	<b>Bibliography .....</b>	<b>41</b>
	<b>Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices .....</b>	<b>43</b>