

# ISO 11979-3:2012-12 (E)

## Ophthalmic implants - Intraocular lenses - Part 3: Mechanical properties and test methods

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
<b>1</b>	<b>Scope .....</b>	<b>1</b>
<b>2</b>	<b>Normative references .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Requirements .....</b>	<b>1</b>
4.1	General .....	1
4.2	Tolerances and dimensions .....	2
4.3	Clearance analysis (anterior chamber lenses only) .....	2
4.4	Compression force .....	2
4.5	Axial displacement in compression .....	2
4.6	Optic decentration .....	3
4.7	Optic tilt .....	3
4.8	Angle of contact .....	3
4.9	Compression force decay .....	3
4.10	Dynamic fatigue durability .....	3
4.11	Surgical manipulation .....	4
4.12	Surface and bulk homogeneity .....	4
<b>5</b>	<b>Recovery of properties following simulated surgical manipulation .....</b>	<b>4</b>
<b>6</b>	<b>Additions for accommodating IOLs (AIOLs) .....</b>	<b>4</b>
<b>Annex A (normative) Measurement of compression force .....</b>		<b>6</b>
<b>Annex B (normative) Measurement of axial displacement in compression .....</b>		<b>9</b>
<b>Annex C (normative) Measurement of optic decentration .....</b>		<b>12</b>
<b>Annex D (normative) Measurement of optic tilt .....</b>		<b>15</b>
<b>Annex E (normative) Measurement of angle of contact .....</b>		<b>19</b>
<b>Annex F (normative) Testing of compression force decay .....</b>		<b>22</b>
<b>Annex G (normative) Testing of dynamic fatigue durability .....</b>		<b>23</b>
<b>Annex H (informative) Measurement of loop pull strength .....</b>		<b>25</b>
<b>Annex I (informative) Clearance analysis .....</b>		<b>27</b>
<b>Annex J (informative) Precision .....</b>		<b>30</b>
<b>Bibliography .....</b>		<b>31</b>