ISO/TS 81060-5:2020-02 (E)

Non-invasive sphygmomanometers - Part 5: Requirements for the repeatability and reproducibility of NIBP simulators for testing of automated non-invasive sphygmomanometers

Contents

Page

Foreword ... **iv**

Introduction ... **v**

1 Scope ... **1**

2 Normative references ... **1**

3 Terms and definitions ... **2**

4 Requirement for technical parameters ... **2**

4.1 Accuracy of the static pressure ... 2

4.1.1 Apparatus ... 2

4.1.2 Requirements ... 2

4.2 Accuracy of the pulse rate ... 4

4.2.1 Apparatus ... 4

4.2.2 Requirement ... 4

4.3 Repeatability of oscillation amplitudes ... 5

4.3.1 Apparatus ... 5

4.3.2 Requirements ... 5

4.4 Reproducibility of oscillation amplitudes ... 5

4.4.1 Apparatus ... 5

4.4.2 Requirements ... 6

4.5 Repeatability and reproducibility of the envelope of the oscillations ... 7

4.6 Repeatability of the shape of oscillations ... 7

4.6.1 ***Procedure*** ... 7

4.6.2 Requirement ... 7

4.7 Reproducibility of the shape of oscillation ... 8

4.7.1 ***Procedure*** ... 8

4.7.2 Requirement ... 8

Annex A (informative) **Particular guidance and rationale** ... **9**

Annex B (informative) **Terminology — Alphabetical index of defined terms** ... **10**

Bibliography ... **11**