

# ISO 7197:2024-07 (E)

## Neurosurgical implants - Sterile, single-use hydrocephalus shunts

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
1	Scope .....	1
2	Normative references .....	1
3	Terms and definitions .....	1
4	General requirements for shunts .....	2
4.1	General .....	2
4.2	Radiopacity .....	2
4.3	Biocompatibility .....	2
4.4	Resistance to leakage .....	2
4.5	Control of the implanted shunt .....	3
4.6	Pressure-flow characteristics of the valve, the components and the pre-assembled shunt .....	3
4.7	Identification of shunts in vivo .....	3
4.8	Ability to withstand overpressure .....	3
4.9	Dynamic breaking strength .....	3
4.10	Behaviour under magnetic resonance imaging conditions .....	3
4.11	Bursting pressure .....	4
5	Specific requirements for components .....	4
5.1	Valves .....	4
5.1.1	Reflux performance of shunts connecting the ventricle to the blood system .....	4
5.1.2	Long-term stability .....	4
5.1.3	Influence of the changed posture of the patient on the valve performance .....	4
5.2	Resistance of tubing and components .....	5
6	Marking and labelling of shunts .....	5
7	Packaging .....	5
8	Information supplied by the manufacturer .....	5
8.1	General .....	5
8.2	Instructions for use .....	5
8.3	Implant card .....	6
Bibliography .....		7