

ISO 13402:2025-03 (E)

Surgical and dental hand instruments - Determination of resistance against autoclaving, corrosion and thermal exposure

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Test methods.....	1
4.1 General.....	1
4.2 Sampling plan.....	1
4.3 Overview of test methods and applicability.....	2
5 Test report.....	2
Annex A (normative) Boiling test in deionized water.....	4
Annex B (normative) Boiling test in 0,9 % NaCl solution.....	6
Annex C (normative) Copper sulfate test for steels with equal or greater than 16 % chromium.....	8
Annex D (normative) Copper sulfate test for martensitic steels with less than 16 % chromium.....	10
Annex E (informative) Test in 0,3 % sodium chloride solution of austenitic steels.....	12
Annex F (informative) Test with citric acid solution for austenitic steels.....	14
Annex G (informative) Thermal test.....	16
Annex H (informative) Autoclave test.....	17
Bibliography.....	18