

# ISO 12487:2026-05 (E)

## Medical electrical equipment - Clinical performance evaluation of clinical thermometers

---

Contents		Page
Foreword		iv
Introduction		v
<b>1</b>	<b>Scope</b>	<b>1</b>
<b>2</b>	<b>Normative references</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions</b>	<b>1</b>
<b>4</b>	<b>General requirements</b>	<b>7</b>
4.1	<i>Clinical investigation</i> methods	7
4.2	Good clinical practice	8
4.3	Status of previous <i>clinical investigations</i>	8
4.4	Disclosure of summary of <i>clinical investigation</i>	8
<b>5</b>	<b><i>Clinical investigation</i></b>	<b>8</b>
5.1	Subject requirements	8
5.1.1	Number	8
5.1.2	Age distribution	8
5.1.3	Body temperature distribution	9
5.1.4	Body site for <i>clinical investigation</i>	9
5.1.5	Exception criteria	10
5.2	<i>Clinical investigation</i> method with a <i>reference thermometer</i>	10
5.2.1	Environmental condition of <i>clinical investigation</i>	10
5.2.2	Subject preparation for body temperature stabilization	10
5.2.3	<i>Reference thermometer</i> measurement	11
5.2.4	<i>Clinical thermometer-under-test</i>	11
5.3	<i>Procedure</i>	11
5.3.1	<i>Clinical thermometers</i> with <i>time prediction</i>	12
5.3.2	<i>Clinical thermometers</i> with <i>site conversion</i>	12
<b>6</b>	<b>Data analysis</b>	<b>13</b>
6.1	Correction of the readings of the <i>reference thermometer</i>	13
6.2	<i>Clinical bias</i> calculation	13
6.3	Standard deviation of the <i>clinical bias</i> calculation	14
6.4	<i>Clinical repeatability</i> calculation	14
<b>7</b>	<b>Acceptance criteria</b>	<b>14</b>
7.1	Clinical bias	14
7.2	Standard deviation of the <i>clinical bias</i>	15
7.3	<i>Clinical repeatability</i>	15
<b>Annex A</b> (informative)	<b>Particular guidance and rationale</b>	<b>16</b>
<b>Annex B</b> (informative)	<b>Guide to disclosure requirements</b>	<b>22</b>
<b>Annex C</b> (informative)	<b>Terminology — alphabetized index of defined terms</b>	<b>23</b>
<b>Bibliography</b>		<b>25</b>