

DIN EN 1060-3:2006-11 (E)

Non-invasive sphygmomanometers - Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

Contents		Page
1	Scope	4
2	Normative references	4
3	Terms and definitions	4
4	Cuff	5
5	Display	5
6	Units	6
7	Requirements	6
7.1	General	6
7.2	Limits of the error of the cuff pressure indication	6
7.3	Effect of voltage variations of the power source variations	6
7.4	Pneumatic system	6
7.5	Environmental performance	7
7.6	Stability of the cuff pressure indication	8
7.7	Pressure indicating device	8
7.8	Signal input and output parts	8
7.9	Overall system accuracy	9
7.10	Alarms	9
7.11	Safety	9
8	Test methods	9
8.1	General	9
8.2	Test methods for the effect of voltage variations of the power source on the cuff pressure reading	10
8.3	Test methods for the effect of voltage variations of the power source on the result of the blood pressure measurement	12
8.4	Test method for the air leakage of the pneumatic system	14
8.5	Test method for the pressure reduction rate	15
8.6	Test method for the rapid exhaust valve	15
8.7	Test method for the zero setting	16
8.8	Test method for the drift of the cuff pressure indication	17
8.9	Test method for the stability of the blood pressure determination	17
8.10	Test method for the stability of cuff pressure indication following prolonged usage	18
8.11	Test method for the effect of external voltages and abnormal connections to the signal input/output part	19
8.12	Test method for the cuff pressure deflation following an aborted measurement	19
9	Information supplied by the manufacturer	19
9.1	General	19
9.2	Instructions for use	20
9.3	Marking of devices	20
Annex A (informative) Rationale		21
Annex B (informative) Bibliography		22
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EC Medical Devices		23