ISO 14937:2009-10 (E)

Sterilization of healt h care products - General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices

Contents

Page

Foreword v

Introduction vi

1 Scope 1
1.1 Inclusions 1
1.2 Exclusions 1

2 Normative references 2

3 Terms and definitions 2

4 Quality management system elements 7
4.1 Documentation 7
4.2 Management responsibility 7
4.3 Product realization 8
4.4 Measurement, analysis and improvement -- Control of non-conforming product 8

5 Sterilizing agent characterization 8
5.1 General 8
5.2 Sterilizing agent 8
5.3 Microbicidal effectiveness 8
5.4 Effects on materials 9
5.5 Safety and the environment 9

6 Process and equipment characterization 9
6.1 General 9
6.2 Process characterization 9
6.3 Equipment characterization 10

7 Product definition 10

8 Process definition 11

9 Validation 12
9.1 General 12
9.2 Installation qualification 12
9.3 Operational qualification 13
9.4 Performance qualification 13
9.5 Review and approval of validation 14

10 Routine monitoring and control 14
11 Product release from sterilization 14
12 Maintaining process effectiveness 15
12.1 General 15
12.2 Recalibration 15
12.3 Maintenance of equipment 15
12.4 Requalification 15
12.5 Assessment of change 15

Annex A (normative) Factors to be considered in selection of microorganisms for demonstrating microbicidal effectiveness 16

Annex B (normative) Approach 1 -- Process definition based on inactivation of the microbial population in its natural state 18

Annex C (normative) Approach 2 -- Process definition based on inactivation of reference microorganisms and knowledge of bioburden 19

Annex D (normative) Approach 3 -- Conservative process definition based on inactivation of reference microorganisms 20

Annex E (informative) Guidance on application of this International Standard 22

Bibliography 36