

ISO 14937:2009-10 (E)

Sterilization of health care products - General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
1.1	Inclusions	1
1.2	Exclusions	1
2	Normative references	2
3	Terms and definitions	2
4	Quality management system elements	7
4.1	Documentation	7
4.2	Management responsibility	7
4.3	Product realization	8
4.4	Measurement, analysis and improvement -- Control of non-conforming product	8
5	Sterilizing agent characterization	8
5.1	General	8
5.2	Sterilizing agent	8
5.3	Microbicidal effectiveness	8
5.4	Effects on materials	9
5.5	Safety and the environment	9
6	Process and equipment characterization	9
6.1	General	9
6.2	Process characterization	9
6.3	Equipment characterization	10
7	Product definition	10
8	Process definition	11
9	Validation	12
9.1	General	12
9.2	Installation qualification	12
9.3	Operational qualification	13
9.4	Performance qualification	13
9.5	Review and approval of validation	14
10	Routine monitoring and control	14
11	Product release from sterilization	14
12	Maintaining process effectiveness	15
12.1	General	15
12.2	Recalibration	15
12.3	Maintenance of equipment	15
12.4	Requalification	15
12.5	Assessment of change	15

Annex A (normative) Factors to be considered in selection of microorganisms for demonstrating microbicidal effectiveness	16
Annex B (normative) Approach 1 -- Process definition based on inactivation of the microbial population in its natural state	18
Annex C (normative) Approach 2 -- Process definition based on inactivation of reference microorganisms and knowledge of bioburden	19
Annex D (normative) Approach 3 -- Conservative process definition based on inactivation of reference microorganisms	20
Annex E (informative) Guidance on application of this International Standard	22
Bibliography	36