ISO/IEEE 11073-10407:2010-05 (E)

Health informatics — Personal health device communication — Part 10407: Device specialization — Blood pressure monitor

Contents

Page

Foreword v

Introduction vii

1. Overview 1

1.1 Scope 1
1.2 Purpose 1
1.3 Context 1

2. Normative references 2

3. Definitions, acronyms, and abbreviations 2

3.1 Definitions 2
3.2 Acronyms and abbreviations 3

4. Introduction to ISO/IEEE 11073 personal health devices 3

4.1 General 3
4.2 Introduction to IEEE 11073-20601 modeling constructs 4

5. Blood pressure monitor device concepts and modalities 5

5.1 General 5
5.2 Systolic and diastolic pressure 5
5.3 Mean arterial pressure 5
5.4 Pulse rate 5

6. Blood pressure monitor domain information model 6

6.1 Overview 6
6.2 Class extensions 6
6.3 Object instance diagram 6
6.4 Types of configuration 7
6.5 Medical device system object 7
6.6 Numeric objects 11
6.7 Real-time sample array objects 13
6.8 Enumeration objects 13
6.9 PM-store objects 15
6.10 Scanner objects 15
6.11 Class extension objects 15
6.12 Blood pressure monitor information model extensibility rules 15

7. Blood pressure monitor service model 15

7.1 General 15
7.2 Object access services 15
7.3 Object access event report services 16

8. Blood pressure monitor communication model 17

8.1 Overview 17
8.2 Communication characteristics 17
8.3 Association procedure 17
8.4 Configuring procedure 19
8.5 Operating procedure 21
8.6 Time synchronization 21

9. Test associations 21

9.1 General 21
9.2 Behavior with standard configuration 22
9.3 Behavior with extended configurations 22

10. Conformance 22

10.1 Applicability 22
10.2 Conformance specification 22
10.3 Levels of conformance 23
10.4 Implementation conformance statements 23

Annex A (informative) Bibliography 28

Annex B (normative) Any additional ASN.1 definitions 29

Annex C (normative) Allocation of identifiers 30

Annex D (informative) Message sequence examples 31

Annex E (informative) Protocol data unit examples 33