

DIN EN ISO 14160:2011-10 (E)

Sterilization of health care products - Liquid chemical sterilizing agents for single-use medical devices utilizing animal tissues and their derivatives - Requirements for characterization, development, validation and routine control of a sterilization process for medical devices (ISO 14160:2011)

Contents		Page
Foreword		3
Introduction		4
1 Scope		6
2 Normative references		7
3 Terms and definitions		7
4 Quality management system elements		10
4.1 Documentation		10
4.2 Management responsibility		11
4.3 Product realization		11
4.4 Measurement, analysis and improvement -- Control of non-conforming products		11
5 Sterilizing agent characterization		11
5.1 General		11
5.2 Sterilizing agent		11
5.3 Microbicidal effectiveness		12
5.4 Effects on materials		12
5.5 Safety and the environment		12
6 Process and equipment characterization		12
6.1 General		12
6.2 Process characterization		13
6.3 Equipment characterization		13
7 Product definition		13
8 Process definition		14
8.1 Purpose		14
8.2 Determination of the inactivation kinetics		14
8.3 Method for neutralization		15
8.4 Safety quality and performance		15
9 Validation		15
9.1 General		15
9.2 Installation qualification		16
9.3 Operational qualification		16
9.4 Performance qualification		16
9.5 Review and approval of validation		18
10 Routine monitoring and control		19
11 Product release from sterilization		21
12 Maintaining process effectiveness		21

12.1	General	21
12.2	Maintenance of equipment	21
12.3	Requalification	21
12.4	Assessment of change	22
Annex A (informative) Guidance for the application of this International Standard		23
Annex B (normative) Determination of lethal rate of the sterilization process		34
Annex C (informative) Flowchart for microbicidal effectiveness (see 5.3), process definition (see Clause 8), and microbiological performance qualification (see 9.4.2)		38
Bibliography		39
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42 EEC on Medical Devices		40