

DIN EN ISO 14160:2011-10 (E)

Sterilization of health care products - Liquid chemical sterilizing agents for single-use medical devices utilizing animal tissues and their derivatives - Requirements for characterization, development, validation and routine control of a sterilization process for medical devices (ISO 14160:2011)

| Contents | | Page |
|--------------------|---|-------------|
| Foreword | | 3 |
| Introduction | | 4 |
| 1 | Scope | 6 |
| 2 | Normative references | 7 |
| 3 | Terms and definitions | 7 |
| 4 | Quality management system elements | 10 |
| 4.1 | Documentation | 10 |
| 4.2 | Management responsibility | 11 |
| 4.3 | Product realization | 11 |
| 4.4 | Measurement, analysis and improvement -- Control of non-conforming products | 11 |
| 5 | Sterilizing agent characterization | 11 |
| 5.1 | General | 11 |
| 5.2 | Sterilizing agent | 11 |
| 5.3 | Microbicidal effectiveness | 12 |
| 5.4 | Effects on materials | 12 |
| 5.5 | Safety and the environment | 12 |
| 6 | Process and equipment characterization | 12 |
| 6.1 | General | 12 |
| 6.2 | Process characterization | 13 |
| 6.3 | Equipment characterization | 13 |
| 7 | Product definition | 13 |
| 8 | Process definition | 14 |
| 8.1 | Purpose | 14 |
| 8.2 | Determination of the inactivation kinetics | 14 |
| 8.3 | Method for neutralization | 15 |
| 8.4 | Safety quality and performance | 15 |
| 9 | Validation | 15 |
| 9.1 | General | 15 |
| 9.2 | Installation qualification | 16 |
| 9.3 | Operational qualification | 16 |
| 9.4 | Performance qualification | 16 |
| 9.5 | Review and approval of validation | 18 |
| 10 | Routine monitoring and control | 19 |
| 11 | Product release from sterilization | 21 |
| 12 | Maintaining process effectiveness | 21 |

| | | |
|---|---------------------------------------|-----------|
| 12.1 | General | 21 |
| 12.2 | Maintenance of equipment | 21 |
| 12.3 | Requalification | 21 |
| 12.4 | Assessment of change | 22 |
| Annex A (informative) Guidance for the application of this International Standard | | 23 |
| Annex B (normative) Determination of lethal rate of the sterilization process | | 34 |
| Annex C (informative) Flowchart for microbicidal effectiveness (see 5.3), process definition (see Clause 8), and microbiological performance qualification (see 9.4.2) | | 38 |
| Bibliography | | 39 |
| Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42 EEC on Medical Devices | | 40 |