DIN EN ISO 11137-2:2012-07 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2012)

Contents

Page

Foreword 4

Introduction 5

1 Scope 6

2 Normative references 6

3 Terms, definitions and abbreviated terms 6
3.1 Terms and definitions 6
3.2 Abbreviated terms 8

4 Definition and maintenance of product families for dose setting, dose substantiation and sterilization dose auditing 9
4.1 General 9
4.2 Defining product families 10
4.3 Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit 10
4.4 Maintaining product families 12
4.5 Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family 12

5 Selection and testing of product for establishing the sterilization dose 12
5.1 Nature of product 12
5.2 Sample item portion (SIP) 13
5.3 Manner of sampling 14
5.4 Microbiological testing 14
5.5 Irradiation 15

6 Methods of dose establishment 15

7 Method 1: dose setting using bioburden information 16
7.1 Rationale 16
7.2 Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for multiple production batches 17
7.3 Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for a single production batch 22
7.4 Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches 24

8 Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor 25
8.1 Rationale 25
8.2 Procedure for Method 2A 26
8.3 Procedure for Method 2B 29

9 Method VDmax -- Substantiation of 25 kGy or 15 kGy as the sterilization dose 33
9.1 Rationale 33
9.2 Procedure for Method VDmax 25 for multiple production batches 34
9.3 Procedure for Method VDmax 25 for a single production batch 39
9.4 Procedure for Method VDmax 15 for multiple production batches 41

9.5 Procedure for Method VDmax 15 for a single production batch 45

10 Sterilization dose audit 48
10.1 Purpose and frequency 48

10.2 Procedure for auditing a sterilization dose established using Method 1, Method 2A or Method 2B 48
10.3 Procedure for auditing a sterilization dose substantiated using Method VDmax 25 or Method VDmax 15 51
10.4 Failure of a sterilization dose audit 55

11 Worked examples 55
11.1 Worked examples for Method 1 55
11.2 Worked examples for Method 2 58
11.3 Worked examples for Method VDmax 68
11.4 Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose 70
11.5 Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose 71
11.6 Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax 25 72

Bibliography 74

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 90/385/EEC on active implantable medicaldevices 76

Annex ZB (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices 77

Annex ZC (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 98/79/EC on in vitro diagnostic medical devices 78