ISO 11608-1:2012-04 (E)

Needle-based injection systems for medical use - Requirements and test methods - Part 1: Needle-based injection systems

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 2

4 Symbols and abbreviated terms 3

5 Requirements 4
5.1 General 4
5.2 System designations 4
5.3 Risk analysis requirements 5
5.4 Uncertaintyofmeasurementandconformancewithspecifications 5
5.5 General design requirements 5

6 Reagent and apparatus 7
6.1 General 7
6.2 Test liquid 7
6.3 Balance 7
6.4 Test surface for free-fall testing 7

7 Determination of dose accuracy 7
7.1 General 7
7.2 Dosing regions 8
7.3 Dose settings 8
7.4 Assessment 9

8 Preparation and operation of NISs 12

9 Test matrix 12

10 Test descriptions 15
10.1 General 15
10.2 Cool, standard and warm atmosphere testing 15
10.3 Last-dose testing (system designations A and C only) 16
10.4 Life-cycle testing (systems designations A and B only) -- Pre-conditioning 16
10.5 Free-fall testing 16
10.6 Dry-heat and cold-storage testing -- Pre-conditioning 18
10.7 Damp-heat testing (system designations A and B only) -- Pre-conditioning 18
10.8 Cyclical testing (system designations A and B only) -- Pre-conditioning 18
10.9 Vibration testing -- Pre-conditioning 18
10.10 Electromagnetic compatibility (EMC) (systems with electronics only) 19

11 Inspection 20
11.1 Visual inspection 20
11.2 Container inspection 20
11.3 Dose accuracy acceptance criteria 20

12 Test report 21

13 Information supplied by the manufacturer 22
13.1 General 22
13.2 Marking 22
13.3 Instructions for use 23

Annex A (informative) Dose replicates, accuracy and testing rationale 25

Annex B (normative) One- and two-sided tolerance limit factors, k 29

Bibliography 40