

DIN EN ISO 18113-5:2013-01 (E)

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing (ISO 18113-5:2009)

| Contents | | Page |
|--|--|-------------|
| Foreword | | 3 |
| Introduction | | 4 |
| 1 | Scope | 5 |
| 2 | Normative references | 5 |
| 3 | Terms and definitions | 6 |
| 4 | Essential requirements | 6 |
| 5 | Labels and marking | 6 |
| 5.1 | General | 6 |
| 5.2 | Identification of the IVD instrument | 6 |
| 6 | Elements of the instructions for use | 7 |
| 7 | Content of the instructions for use | 7 |
| 7.1 | Manufacturer | 7 |
| 7.2 | Identification of the IVD instrument | 7 |
| 7.3 | Intended use | 8 |
| 7.4 | Storage and handling | 8 |
| 7.5 | Warnings and precautions | 8 |
| 7.6 | Instrument installation | 8 |
| 7.7 | Principles of measurement | 9 |
| 7.8 | Performance of the IVD instrument | 9 |
| 7.9 | Limitations of use | 9 |
| 7.10 | Preparation prior to operation | 9 |
| 7.11 | Operating procedure | 9 |
| 7.12 | Control procedure | 10 |
| 7.13 | Reading of examination results | 10 |
| 7.14 | Special functions | 10 |
| 7.15 | Shut-down procedure | 10 |
| 7.16 | Disposal information | 10 |
| 7.17 | Maintenance | 11 |
| 7.18 | Troubleshooting | 11 |
| 7.19 | Follow-up action | 11 |
| Bibliography | | 12 |
| Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of the EU Directive 98/79/EC on "in vitro Diagnostic Medical Devices" | | 13 |