

# ISO/TS 13004:2013-05 (E)

## Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD<(Inde x)max><(hoch)SD>

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
<b>1</b>	<b>Scope .....</b>	<b>1</b>
1.1	Inclusions .....	1
1.2	Exclusions .....	1
1.3	Application .....	1
<b>2</b>	<b>Normative references .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing .....</b>	<b>5</b>
4.1	General .....	5
4.2	Defining product families .....	5
4.3	Designation of product to represent a product family .....	6
4.4	Maintaining product families .....	7
4.5	Consequence of failure of sterilization dose substantiation or sterilization dose audit .....	8
<b>5</b>	<b>Selection and testing of product for substantiating and auditing a selected sterilization dose .....</b>	<b>8</b>
5.1	Nature of product .....	8
5.2	Sample item portion (SIP) .....	9
5.3	Manner of sampling .....	10
5.4	Microbiological testing .....	11
5.5	Irradiation .....	11
<b>6</b>	<b>Method VD SD max -- Substantiation of a selected sterilization dose of 17,5, 20, 22,5, 27,5, 30, 32,5, or 35 kGy .....</b>	<b>11</b>
6.1	Rationale .....	11
6.2	Procedure for Method VD SD max for multiple production batches .....	12
6.3	Procedure for Method VD SD max for a single production batch .....	17
<b>7</b>	<b>Maintaining process effectiveness .....</b>	<b>21</b>
7.1	General .....	21
7.2	Determination of bioburden .....	22
7.3	Sterilization dose audit .....	22
<b>8</b>	<b>Tables of values for SIP equal to 1,0 VDmaxSD, SIP dose reduction factor and augmentation dose corresponding to applicable values of average bioburden for selected sterilization doses of 17,5, 20, 22,5, 27,5, 30, 32,5 and 35 kGy .....</b>	<b>27</b>
<b>9</b>	<b>Worked examples .....</b>	<b>53</b>
9.1	Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0) .....	53
9.2	Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0) .....	54
9.3	Sterilization dose audit for a sterilization dose substantiated using Method VDmax22,5, the findings from which necessitated augmentation of the sterilization dose .....	55
Bibliography .....		57