

ISO 15197:2013-05 (E)

In vitro diagnostic test systems - Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	2
4	Design and development	8
4.1	General requirements	8
4.2	Metrological traceability	8
4.3	Safety and risk management	9
4.4	Ergonomics and human factors	10
4.5	User verification requirements	10
5	Safety and reliability testing	10
5.1	General requirements	10
5.2	Protection against electric shock	11
5.3	Protection against mechanical hazards	11
5.4	Electromagnetic compatibility	11
5.5	Resistance to heat	11
5.6	Resistance to moisture and liquids	11
5.7	Protection against liberated gases, explosion and implosion	12
5.8	Meter components	12
5.9	Performance test	12
5.10	Mechanical resistance to vibration and shock	12
5.11	Equipment temperature exposure limits for storage	13
5.12	Equipment humidity exposure limits for storage	13
6	Analytical performance evaluation	14
6.1	General requirements	14
6.2	Measurement precision	16
6.3	System accuracy	19
6.4	Influence quantities	25
6.5	Stability of reagents and materials	30
7	Information supplied by the manufacturer	30
7.1	General requirements	30
7.2	Performance characteristics	31
7.3	Options for supplying instructions for use	31
8	User performance evaluation	31
8.1	General requirements	31
8.2	Acceptance criteria and evaluation of results	32
8.3	Selection and preparation of subjects	32
8.4	Execution of study protocol	32
8.5	Glucose reference values	33
8.6	Human factors	33
8.7	Data analysis and presentation of results	33

8.8	Evaluation of instructions for use	34
	Annex A (informative) Possible interfering substances	35
	Annex B (informative) Traceability chain	36
	Annex C (informative) Rationale for the analytical performance requirements	38
	Bibliography	45