

ISO 10555-1:2013-06 (E)

Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements

Contents		Page
Foreword		iv
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Requirements	4
4.1	General	4
4.2	Radio-detectability	4
4.3	Biocompatibility	4
4.4	Surface	4
4.5	Corrosion resistance	4
4.6	Peak tensile force	4
4.7	Freedom from leakage	5
4.8	Hubs	5
4.9	Flowrate	5
4.10	Power injection	5
4.11	Side holes	5
4.12	Distal tip	5
5	Designation of nominal size	5
5.1	Outside diameter	5
5.2	Nominal effective length	6
6	Information to be supplied by the manufacturer	6
6.1	General	6
6.2	Marking on the device and/or primary packaging	6
6.3	Instructions for use	7
6.4	Marking on the secondary packaging	7
Annex A (normative) Test method for corrosion resistance		8
Annex B (normative) Method for determining peak tensile force		9
Annex C (normative) Test method for liquid leakage under pressure		11
Annex D (normative) Test method for air leakage into hub assembly during aspiration		13
Annex E (normative) Determination of flowrate through catheter		15
Annex F (normative) Test for burst pressure under static conditions		17
Annex G (normative) Power injection test for flowrate and device pressure (only for products indicated for power injection)		19
Annex H (informative) Units of measurement systems other than those specified in this part of Bibliography		24