

DIN EN 1422:2014-08 (E)

Sterilizers for medical purposes - Ethylene oxide sterilizers - Requirements and test methods

Contents	Page
Foreword	4
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
4 Technical requirements	12
4.1 General	12
4.1.1 Risk control and usability	12
4.1.2 Materials	13
4.2 Sterilizer chamber	13
4.2.1 Chamber size	13
4.2.2 Doors, closures and interlocks of the sterilizer chamber	13
4.2.3 Test connections	14
4.3 Design and construction	15
4.3.1 General	15
4.3.2 EO vaporizers	15
4.3.3 Pipework and fittings	15
4.3.4 Evacuation system	15
4.3.5 Control valves	16
4.3.6 Thermal insulation	16
4.3.7 Electrical and mechanical safety	16
4.3.8 Air or inert gas filter	16
4.3.9 Emission control	16
4.3.10 Framework and panelling	17
4.3.11 Loading equipment	17
4.3.12 Transport	17
4.4 Indicating, measuring, and recording instruments	17
4.4.1 General	17
4.4.2 Temperature sensor	18
4.4.3 Temperature indicating instruments	18
4.4.4 Pressure sensors	19
4.4.5 Timers and time indicating instruments	19
4.4.6 Sterilizing cycle counter	19
4.4.7 Relative humidity (RH) sensors	19
4.4.8 Ethylene Oxide (EO) concentration-measurement	19
4.4.9 Recording instruments	20
4.4.10 Indicating instruments	21
5 Process control	22
5.1 General	22
5.2 Software verification and validation	23
5.3 Sterilization cycle and automatic control	23
5.3.1 Automatic control	23
5.3.2 Sterilization cycle	24
5.4 Override of automatic control	27
5.5 Fault	27

6	Performance requirements	28
6.1	Sterilizing performance	28
6.1.1	Loading configuration	28
6.1.2	Physical parameters	28
6.1.3	Microbiological efficacy	28
6.2	EO removal (flushing)	29
6.3	Aeration	29
7	Sound power	29
8	Packaging, marking and labelling	29
9	Information to be supplied by the manufacturer	30
10	Service and local environment	32
10.1	General	32
10.2	Electricity	33
10.3	Sterilant	33
10.4	Circulation systems	33
10.5	Steam	33
10.6	Water	34
10.7	Air and inert gasses	34
10.8	Drainage and discharges	34
10.9	Ventilation and environment	34
10.10	Lighting	34
	Annex A (normative) Test instrumentation	35
	Annex B (normative) Leak test cycle	36
	Annex C (normative) Sterilizer chamber profile testing	37
C.1	Sterilizer chamber internal surfaces	37
C.2	Empty sterilizer chamber	37
	Annex D (normative) Microbiological test for EO sterilizers	38
D.1	General	38
D.2	Test equipment	38
D.3	Procedure	39
D.4	Interpretation of results	40
	Annex E (informative) Environmental aspects	41
E.1	Environmental aspects regarding the life cycle of EO sterilizers	41
E.2	EO (brief description)	41
E.3	Environmental impact	41
	Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices	44
	Bibliography	48