DIN EN 13718-1:2014-12 (E)

Medical vehicles and their equipment - Air ambulances - Part 1: Requirement s for medical devices used in air ambulances

Contents

Page

Foreword 3

Introduction 4

1 Scope 5

2 Normative references 5

3 Terms and definitions 6

4 Requirements for medical devices for air ambulances 6
4.1 General 6
4.2 Patient and personnel safety 7
4.3 User interface 7
4.4 Environmental conditions and performance of medical devices intended for use in air ambulances 7
4.4.1 Functional temperature range 7
4.4.2 Humidity 8
4.4.3 Variable atmospheric pressures 8
4.5 Electrically-powered medical devices 8
4.5.1 General 8
4.5.2 Medical devices with 12 V DC power input 8
4.5.3 Medical devices with 24 V DC power input 8
4.5.4 Medical devices with 230 V AC power input 9
4.5.5 Short time voltage drop 9
4.5.6 Internal electrical power source 9
4.5.7 Electromagnetic interference of medical devices 9
4.6 Medical gas supply 9
4.6.1 General 9
4.6.2 Gas leakage 9
4.6.3 Pressure regulators and flow metering devices 9
4.6.4 Pneumatic power 10
4.6.5 Cylinder valves 10
4.6.6 Low pressure hose assemblies 10
4.7 Mechanical strength 10
4.7.1 General 10
4.7.2 Vibration and bump 10
4.7.3 Free fall 10
4.8 Fixation of medical devices in air ambulances 10
4.9 Fire resistance 10
4.10 Information to be supplied by the manufacturer 11

5 Test methods 11
5.1 General 11
5.2 Ambient conditions 11
5.3 Test method for durability of markings and colour coding 11
5.4 Free fall 11

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices 13

Bibliography 15