

ISO/TR 13570-2:2014-06 (E)

Wheelchairs - Part 2: Typical values and recommended limits of dimensions, mass and manoeuvring space as determined in ISO 7176-5

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Wheelchair groups	1
4.1	General	1
4.2	Wheelchairs with handrims	2
4.3	Electrically powered wheelchairs of class A	2
4.4	Electrically powered wheelchairs of class B	2
4.5	Electrically powered wheelchairs of class C	2
4.6	Electrically powered wheelchairs (scooter design)	2
5	Typical values and recommended limits for required measurements	2
5.1	General	2
5.2	Full overall length	2
5.3	Overall width	3
5.4	Handgrip height	4
5.5	Stowage length	5
5.6	Stowage width	6
5.7	Stowage height	7
5.8	Rising	8
5.9	Total mass	8
5.10	Mass of heaviest part	9
5.11	Pivot width	10
5.12	Reversing width	11
5.13	Turning diameter	12
5.14	Ground clearance	13
5.15	Required width of angled corridor	14
5.16	Required doorway entry depth	15
5.17	Required corridor width for side opening	16
6	Typical values and recommended limits for Technical dimensions	17
6.1	Reduced overall length	17
6.2	Overall height	18
6.3	Radial wheel deviation for mass group I, II, and III	18
6.4	Lateral wheel deviation for mass group I, II, and III	19
6.5	Radial handrim deviation for mass I, II, and III	19
6.6	Lateral handrim deviation for mass I, II, and III	20
6.7	Full occupied length	20
6.8	Reduced occupied length	21
6.9	Occupied width	22
6.10	Occupied height	23
6.11	Ramp transition angle	23
6.12	Wheelbase	24
6.13	Rear wheel track	26

6.14	Front wheel track	27
6.15	Camber	28
6.16	Toe of occupant mass group I, II, and III	29
6.17	Skew of occupant mass group I, II, and III	30
6.18	Castor rake of occupant mass group I, II, and III	31
6.19	Castor cant for occupant mass group I, II, and III	32
6.20	Castor trail for occupant mass group I, II, and III	32
6.21	Castor wheel misalignment of occupant mass group I, II, and III	33