ISO 3826-4:2015-08 (E)

Plastics collapsible containers for human blood and blood components - Part 4: Aphaeresis blood bag systems with integrated features

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 2

4 Dimensions 4

5 Design 10
5.1 Leucocyte filter 10
5.2 Pilot samples 10
5.3 Access line needle and return line needle 11
5.4 Needle stick protection device 11
5.5 Pre-collection sampling device 11
5.6 Red blood cell storage bag 11
5.7 Plasma storage bag 12
5.8 Platelet storage bag 12
5.9 Post-collection sampling device 12
5.10 Collection and transfer tube(s) 12
5.11 Outlet port(s) 13
5.12 Suspension 13

6 Requirements 13
6.1 General 13
6.2 Physical requirements 14
6.2.1 Conditions of manufacture 14
6.2.2 Sterilization 14
6.2.3 Transparency 14
6.2.4 Coloration 14
6.2.5 Thermal stability 14
6.2.6 Water vapour transmission for plastics containers prefilled with storage solution or anticoagulant 14
6.2.7 Resistance to leakage 15
6.2.8 Insertion force 15
6.2.9 Pull force 15
6.2.10 Leakage after closure piercing 15
6.2.11 Particulate contamination 15
6.3 Chemical requirements 16
6.3.1 Requirements for the raw container or sheeting 16
6.3.2 Requirements for the test fluid 16
6.4 Biological requirements 17
6.4.1 General 17
6.4.2 Impermeability for microorganisms 17
6.4.3 Compatibility 17

7 Packaging 17
7.1 General 17

7.2 Shelf-life 17
7.3 Over-package materials 17
7.4 Over-package sealing 17
7.5 Over-package strength 18
7.6 Arrangement of components in the over-package 18

8 Labelling 18
8.1 General 18
8.2 Label on plastics containers 18

8.3 Label on over-package 18
8.4 Package insert or instructions for use 19
8.5 Label on shipping box 19
8.6 Label requirements 20

9 Anticoagulant and/or preservative solution 20

Annex A (normative) Chemical tests 21

Annex B (normative) Physical tests 26

Annex C (normative) Biological tests 28

Bibliography 31