ISO 8537:2016-03 (E)

Sterile single-use syringes, with or without needle, for insulin

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 2

4 Types of syringes 5

5 Requirements 5
5.1 General requirements 5
5.2 Material selection 6
5.3 Colour coding 6
5.4 Extraneous matter 7
5.4.1 General 7
5.4.2 Limits for acidity or alkalinity 7
5.4.3 Limits for extractable metals 7
5.5 Lubrication 7
5.5.1 Lubrication of syringes 7
5.5.2 Lubrication of needle tube 8
5.6 Dimensions 8
5.6.1 Barrel and plunger stopper 8
5.6.2 Finger grips 8
5.7 Plunger/plunger stopper 8
5.7.1 General 8
5.7.2 Fit of plunger stopper in barrel 8
5.8 Nozzle 8
5.8.1 Conical fitting 8
5.8.2 Position of nozzle on end of barrel 8
5.9 Needle tubing and needles 9
5.9.1 Needles for syringe types 3 and 4 9
5.9.2 Needle tubing for syringe types 5, 6, 7 and 8 9
5.9.3 Bond between hub and needle tube 9
5.10 Standard test environmental conditions 9
5.11 Performance of assembled syringe 9
5.11.1 Dead space 9
5.11.2 Freedom from leakage at needle 10
5.11.3 Freedom from leakage past plunger stopper 10

6 Packaging 10
6.1 Unit packaging and self-contained syringe units 10
6.1.1 General 10
6.1.2 Unit packaging providing sterile barrier syringes (types 1, 3, 5 and 7) 10
6.1.3 Self-contained syringes with sterile interiors (types 2, 4, 6 and 8) 10
6.2 Multiple-unit packaging (for syringe types 2, 4, 6 and 8) 11
6.3 User packaging 11

7 Information supplied by the manufacturer 11
7.1 General 11

7.2 Syringes 11
7.2.1 General 11
7.2.2 Additional marking for self-contained syringes (syringe types 2, 4, 6 and 8) 12
7.3 Unit packaging (for syringe types 1, 3, 5 and 7) 12
7.4 Multiple unit packs (syringe types 2, 4, 6 and 8) 12
7.5 User packaging 13
7.6 Storage container 14
7.7 Transport wrapping 14

Annex A (normative) Fluid for determination of acidity/alkalinity and extractable metals 15

Annex B (normative) Test method for air leakage past syringe piston during aspiration and for separation of rubber stopper and plunger 16

Annex C (normative) Test method for determination of forces required to operate piston 18

Annex D (normative) Test method for determination of dead space 20

Annex E (normative) Test method for liquid leakage at syringe piston and syringe nozzle/ hub or needle/barrel unions during compression 21

Annex F (normative) Test method for air leakage past nozzle/hub or needle/barrel unions during aspiration 23

Annex G (normative) Preparation of extract for test for pyrogenicity and toxicity 24

Annex H (normative) Syringe sizes and graduated scales 25

Bibliography 27