

# ISO 11040-8:2016-11 (E)

## Prefilled syringes - Part 8: Requirements and test methods for finished prefilled syringes

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
1	Scope .....	1
2	Normative references .....	1
3	Terms and definitions .....	2
4	User requirements .....	2
4.1	Definition of intended use .....	2
4.2	Risk management .....	3
4.3	Application of usability engineering .....	3
5	System characterization .....	3
5.1	Critical dimensions .....	3
5.2	Description of components and materials .....	4
5.2.1	General .....	4
5.2.2	Barrel .....	4
5.2.3	Plunger stoppers .....	5
5.2.4	Additional components .....	5
5.3	Description of the content of the finished prefilled syringe .....	5
6	Performance requirements .....	5
6.1	General .....	5
6.2	Break loose and extrusion forces .....	5
6.3	Burst resistance .....	6
6.4	Break resistance .....	6
6.5	Closure system forces and torques .....	6
6.6	Connectivity with fluid path connectors .....	6
6.7	Residual volume .....	6
6.8	Needle penetration force .....	6
6.9	Needle pull-out force .....	6
6.10	Sharps injury protection requirements .....	6
6.11	Liquid leakage beyond plunger .....	7
6.12	Markings .....	7
7	Pharmaceutical requirements .....	7
7.1	General .....	7
7.2	Drug-container interaction .....	7
7.3	Biological requirements .....	7
7.4	Container closure integrity .....	7
7.5	Deliverable volume .....	8
7.6	Particles (visible and subvisible) .....	8
8	Documentation .....	8
Bibliography .....		9