DIN EN 13060:2019-02 (E)

Small steam sterilizers (includes Amendment :2018)

Contents

Page

European foreword 4

Introduction 6

1 Scope 7

2 Normative references 7

3 Terms and definitions 8

4 General technical requirements 16
4.1 Dimensions 16
4.2 Materials 16
4.3 Design and construction 17
4.4 Instrumentation, indication and registration devices 18
4.5 Control systems 27
4.6 Process 29
4.7 Services and local environment 30
4.8 Information to be provided 31
4.9 Marking 35
4.10 Accessories 36

5 Performance requirements 36
5.1 General 36
5.2 Air leakage rate 37
5.3 Attainment of the sterilization conditions 37
5.4 Product compatibility 37
5.5 Drying 38
5.6 Microbicidal efficacy 38
5.7 Non-condensable gases 38

6 Safety, risk control and usability 38
6.1 General requirements 38
6.2 Requirements for EMC 39
6.3 Requirements for pressure equipment 39
6.4 Requirements for risk control 40

7 Categories of tests 40
7.1 General 40
7.2 Type tests 41
7.3 Works test 41
7.4 Installation tests 42

8 Test equipment 42
8.1 General 42
8.2 Temperature sensors 42
8.3 Thermometric recording instrument 43
8.4 Pressure measurement and recording instrument 43
8.5 Test equipment for the performance of the air leakage test 44
8.6 Porous load 45
8.7 Solid load, unwrapped 47
8.8 Solid load, single wrapped 47

8.9 Solid load, double wrapped 47
8.10 Process challenge device (PCD) and chemical indicator for narrow lumen 48
8.11 Process challenge device and chemical indicator for simple hollow item 48
8.12 Balance for load dryness test 49

9 Test programme 49

10 Test methods 51
10.1 General requirements on technical tests 51
10.2 Air leakage test 51
10.3 Dynamic sterilizer chamber pressure test 53
10.4 Empty chamber test 54
10.5 Solid load test 54
10.6 Narrow lumen test 54
10.7 Simple hollow item test 55
10.8 Small porous load test 56
10.9 Full porous load test (single and double wrapped) 57
10.10 Small porous items test (single and double wrapped) 58
10.11 Solid load dryness test 58
10.12 Porous load dryness test (small and full, single and double wrapped) 59
10.13 Small porous items dryness test (single and double wrapped) 60
10.14 Non-condensable gases test 60
10.15 Microbiological test for solid loads 64
10.16 Microbiological test for narrow lumens 64
10.17 Microbiological test for simple hollow item 64
10.18 Microbiological test for small porous loads 65
10.19 Microbiological test for full porous loads 65
10.20 Microbiological test for small porous items 66

Annex A (informative) Clarification of the definition of narrow lumens and simple hollow items (see 3.18 and 3.30) 67

Annex B (informative) Process evaluation system 69

Annex C (informative) Suggested maximum limits of contaminants in and specification for water for steam sterilization 70

Annex D (informative) Example of a table to be supplied with pre-purchase documentation and with the instructions for use 71

Annex E (informative) Load support systems 72

Annex F (informative) Rationale for the tests 73

Annex G (informative) Example of a process challenge device for narrow lumen 76

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42 EEC on medical devices 77

Bibliography 82