DIN EN ISO 20697:2019-04 (E)

Sterile drainage catheters and accessory devices for single use (ISO 20697:2018, Corrected version 2018-09)

Contents

Page

EN ISO 20697:2018 (E) European foreword 4

Foreword 5

Introduction 6

1 Scope 7

2 Normative references 7

3 Terms and definitions 7

4 Intended performance 9

5 General requirements 9
5.1 Risk management 9
5.2 Biocompatibility 10
5.3 Detectability 10
5.4 Surface finish 10
5.5 Size designation 10
5.5.1 General 10
5.5.2 Outer diameter 10
5.5.3 Effective length 10
5.5.4 Nominal balloon inflation volume 11
5.6 Connector 11
5.7 MRI compatibility 12
5.8 Sterilization 12

6 Specific requirements 12
6.1 Kink stability 12
6.2 Corrosion resistance 12
6.3 Resistance to deformation 12
6.4 Peak tensile force 12
6.4.1 Connections 12
6.4.2 Drainage catheters and other accessory devices 13
6.5 Impact resistance 13
6.6 Flow rate 13
6.7 Retention strength 13
6.8 Balloon safety 13
6.9 Drainage catheter inflation lumen integrity and volume maintenance 14
6.9.1 General 14
6.9.2 Compliant balloon 14
6.9.3 Non-compliant balloon 14
6.10 Inflated balloon resistance to traction 14
6.11 Freedom from leakage during aspiration or vacuum 14

7 Information supplied by the manufacturer 15
7.1 General 15
7.2 Marking on the device and/or packaging 15
7.3 Instructions for use 15

Annex A (informative) Test method for determining kink stability **17**
Annex B (normative) Test method for corrosion resistance **19**
Annex C (normative) Test method for resistance to deformation by suction **20**
Annex D (normative) Test method for determining peak tensile force of connections **21**
Annex E (normative) Test method for determining peak tensile force of drainage catheter **22**
Annex F (normative) Test method for impact resistance of collection device **24**
Annex G (normative) Test method for determination of flow rate through drainage catheter **26**
Annex H (informative) Test method for retention strength **28**
Annex I (normative) Test method for determining balloon safety **30**
Annex J (normative) Test method for determining inflation lumen leakage and/or function and/or balloon deflation (drainage catheter with compliant balloon) **33**
Annex K (normative) Test method for determining balloon size and deflation reliability (drainage catheter with non-compliant balloon) **35**
Annex L (normative) Test method for determining inflated balloon resistance to traction **37**
Annex M (normative) Test method for resistance to leakage during aspiration or vacuum **41**
Bibliography **42**