DIN EN ISO 20695:2020-10 (E)

Enteral feeding systems - Design and testing (ISO 20695:2020)

Contents

Page

European foreword ... **4**

Foreword ... **5**

Introduction ... **6**

1 Scope ... **7**

2 Normative references ... **7**

3 Terms and definitions ... **8**

4 General requirements ... **9**
4.1 General ... 9
4.2 Risk management ... 9
4.3 Usability ... 9
4.4 Test methods ... 9
4.5 Materials ... 9
4.6 Cleaning and disinfection ... 10
4.7 Sterility ... 10
4.8 Packaging ... 10
4.9 Biological safety ... 10
4.10 Corrosion resistance ... 10
4.11 Surface finish ... 10
4.12 Information supplied by the manufacturer ... 10
4.12.1 Marking ... 10
4.12.2 Symbols ... 11
4.12.3 Labelling ... 11
4.12.4 Instructions for use ... 11

5 Additional requirements for enteral giving sets and enteral extension sets ... **12**
5.1 General ... 12
5.2 Inlet ports ... 12
5.3 Outlet ports ... 12
5.4 Access ports ... 13
5.5 Tensile strength ... 13
5.6 Leakage ... 13
5.7 Additional information provided by the manufacturer ... 13

6 Additional requirements for enteral syringes ... **15**
6.1 General ... 15
6.2 Outlet port ... 15
6.3 Enteral syringe requirements ... 15
6.4 Enteral Syringe dose accuracy requirement ... 16
6.5 Marking ... 16

7 Additional requirements for enteral feeding catheters ... **16**
7.1 General ... 16
7.2 Access ports ... 17
7.3 Tensile properties ... 17
7.3.1 Enteral feeding catheters designed for use without an integral introducer system ... 17
7.3.2 Enteral feeding catheters with an integral introducer system ... 17
7.4 Leakage properties ... 17
7.5 Flow rate ... 18
7.6 Enteral feeding catheter designated size ... 18

7.7 Requirements for enteral feeding catheters with retention balloons 18
7.7.1 Balloon burst volume 18
7.7.2 Balloon recommended inflation volume 18
7.7.3 Balloon inflation system performance 18
7.7.4 Balloon concentricity 19
7.7.5 Balloon integrity in simulated gastric fluid 19
7.8 Detectability 19
7.8.1 General 19
7.8.2 Catheters designed to be radiopaque 19
7.8.3 MRI compatibility 19
7.9 Marking 19
8 Additional requirements for enteral accessories 19
Annex A (informative) **Rationale and guidance 21**
Annex B (normative) **Test method for corrosion resistance of metallic components 24**
Annex C (normative) **Test method for tensile properties 26**
Annex D (normative) **Test method for resistance to liquid leakage under pressure 28**
Annex E (normative) **Test method for determining the flow rate 30**
Annex F (normative) **Test method for determining the designated outer diameter of enteral feeding catheters 33**
Annex G (normative) **Test method for determining balloon burst volume 36**
Annex H (normative) **Test method for determining balloon inflation system performance 38**
Annex I (normative) **Test method for determining balloon concentricity 40**
Annex J (normative) **Test method for determining balloon integrity in simulated gastric fluid 42**
Annex K (informative) **Example of an alternative enteral syringe tip 44**
Bibliography 48