

DIN EN ISO 20696:2020-10 (E)

Sterile urethral catheters for single use (ISO 20696:2018, Corrected version 2019-12)

Contents	Page
European foreword	4
Foreword	5
Introduction	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	7
4 Intended performance	9
5 General requirements	9
5.1 Risk management	9
5.2 Biocompatibility	9
5.3 Detectability	9
5.4 Surface finish	9
5.5 Size designation	9
5.5.1 General	9
5.5.2 Outer diameter	9
5.5.3 Effective shaft lengths	10
5.6 MRI compatibility	11
5.7 Connector	11
5.8 Sterilization	11
6 Specific requirements	11
6.1 Strength	11
6.2 Connector security	11
6.3 Balloon safety	11
6.4 Catheter inflation lumen integrity and volume maintenance	11
6.4.1 General	11
6.4.2 Compliant balloon	12
6.4.3 Non-compliant balloon	12
6.5 Flow rate	12
6.6 Corrosion resistance	13
6.7 Kink stability	13
6.8 Peak tensile force	13
6.9 Inflated balloon resistance to traction	13
7 Information to be supplied by the manufacturer	14
7.1 General	14
7.2 Marking on the device and/or packaging	14
7.3 Instructions for use	14
Annex A (normative) Test method for determining the strength of the catheter	15
Annex B (normative) Test method for determining the security of fit of the drainage funnel	19
Annex C (normative) Test method for determining balloon safety	21
Annex D (normative) Test method for determining inflation lumen leakage and/or function and/or balloon deflation (catheter with compliant balloon)	23
Annex E (normative) Test method for determination of flow rate through catheter	25
Annex F (normative) Test method for corrosion resistance	27
Annex G (informative) Test method for determining kink stability	28

Annex H (normative) Test method for determining peak tensile force of urethral catheter	30
Annex I (normative) Test method for determining balloon size deflation reliability (catheter with non-compliant balloon)	32
Annex J (normative) Test method for determining inflated balloon resistance to traction	34
Bibliography	38