

# DIN EN ISO 11137-2:2023-08 (E)

## Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2013 + Amd 1:2022) (includes Amendment A1:2023)

---

<b>Contents</b>		<b>Page</b>
European foreword .....		4
European foreword to Amendment .....		5
Foreword to Amendment .....		6
Introduction .....		7
<b>1</b>	<b>Scope .....</b>	<b>8</b>
<b>2</b>	<b>Normative references .....</b>	<b>8</b>
<b>3</b>	<b>Terms, definitions, and abbreviated terms .....</b>	<b>8</b>
<b>3.1</b>	<b>Terms and definitions .....</b>	<b>8</b>
<b>3.2</b>	<b>Abbreviated terms .....</b>	<b>10</b>
<b>4</b>	<b>Definition and maintenance of product families for dose setting, dose substantiation, and sterilization dose auditing .....</b>	<b>11</b>
<b>4.1</b>	<b>General .....</b>	<b>11</b>
<b>4.2</b>	<b>Defining product families .....</b>	<b>11</b>
<b>4.3</b>	<b>Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit .....</b>	<b>12</b>
<b>4.4</b>	<b>Maintaining product families .....</b>	<b>13</b>
<b>4.5</b>	<b>Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family .....</b>	<b>14</b>
<b>5</b>	<b>Selection and testing of product for establishing the sterilization dose .....</b>	<b>14</b>
<b>5.1</b>	<b>Nature of product .....</b>	<b>14</b>
<b>5.2</b>	<b>Sample item portion (SIP) .....</b>	<b>15</b>
<b>5.3</b>	<b>Manner of sampling .....</b>	<b>16</b>
<b>5.4</b>	<b>Microbiological testing .....</b>	<b>16</b>
<b>5.5</b>	<b>Irradiation .....</b>	<b>16</b>
<b>6</b>	<b>Methods of dose establishment .....</b>	<b>16</b>
<b>7</b>	<b>Method 1: dose setting using bioburden information .....</b>	<b>17</b>
<b>7.1</b>	<b>Rationale .....</b>	<b>17</b>
<b>7.2</b>	<b>Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for multiple production batches .....</b>	<b>18</b>
<b>7.3</b>	<b>Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for a single production batch .....</b>	<b>24</b>
<b>7.4</b>	<b>Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches .....</b>	<b>26</b>
<b>8</b>	<b>Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor .....</b>	<b>27</b>
<b>8.1</b>	<b>Rationale .....</b>	<b>27</b>
<b>8.2</b>	<b>Procedure for Method 2A .....</b>	<b>28</b>
<b>8.3</b>	<b>Procedure for Method 2B .....</b>	<b>31</b>
<b>9</b>	<b>Method VD<sub>max</sub> -- Substantiation of a selected sterilization dose .....</b>	<b>35</b>
<b>9.1</b>	<b>Selected doses and rationale .....</b>	<b>35</b>
<b>9.2</b>	<b>Procedure for Method VD<sub>max25</sub> for multiple production batches .....</b>	<b>36</b>
<b>9.3</b>	<b>Procedure for Method VD<sub>max25</sub> for a single production batch .....</b>	<b>41</b>
<b>9.4</b>	<b>Procedure for Method VD<sub>max15</sub> for multiple production batches .....</b>	<b>44</b>
<b>9.5</b>	<b>Procedure for Method VD<sub>max15</sub> for a single production batch .....</b>	<b>47</b>

<b>10</b>	<b>Sterilization dose audit .....</b>	<b>50</b>
<b>10.1</b>	<b>Purpose and frequency .....</b>	<b>50</b>
<b>10.2</b>	<b>Procedure for auditing a sterilization dose established using Method 1, Method 2A, or Method 2B .....</b>	<b>50</b>
<b>10.3</b>	<b>Procedure for auditing a sterilization dose substantiated using Method VDmax25 or Method VDmax15 .....</b>	<b>53</b>
<b>10.4</b>	<b>Failure of a sterilization dose audit .....</b>	<b>59</b>
<b>11</b>	<b>Worked examples .....</b>	<b>59</b>
<b>11.1</b>	<b>Worked examples for Method 1 .....</b>	<b>59</b>
<b>11.2</b>	<b>Worked examples for Method 2 .....</b>	<b>61</b>
<b>11.3</b>	<b>Worked examples for Method VDmax .....</b>	<b>69</b>
<b>11.4</b>	<b>Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose .....</b>	<b>71</b>
<b>11.5</b>	<b>Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose .....</b>	<b>71</b>
<b>11.6</b>	<b>Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax25 .....</b>	<b>72</b>
<b>Annex ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered .....</b>		<b>74</b>
<b>Annex ZB (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/746 aimed to be covered .....</b>		<b>77</b>
<b>Bibliography .....</b>		<b>79</b>