

DIN EN ISO 18778:2023-02 (E)

Respiratory equipment - Particular requirements for basic safety and essential performance of infant cardiorespiratory monitors (ISO 18778:2022)

Contents

Page

European foreword	5
Foreword	6
Introduction	8
1 Scope	9
2 Normative references	9
3 Terms and definitions	10
4 General requirements	14
4.1 General	14
4.2 Essential performance	14
4.3 ME equipment or ME system parts that contact the patient	14
4.4 Single fault condition for ME equipment	14
5 General requirements for testing of ME equipment	14
5.1 General	14
5.2 Infant cardiorespiratory monitor testing errors	14
6 Classification of ME equipment and ME systems	15
6.1 General	15
6.2 Additional requirements for classification of ME equipment and ME systems	15
7 ME equipment identification, marking and documents	15
7.1 General	15
7.2 Information to be supplied by the manufacturer	15
7.3 Additional requirements for accessories	15
7.4 Additional requirements for marking on the outside of ME equipment or ME equipment parts	16
7.5 General instructions for use	16
7.6 Additional requirements for warnings and safety notices	16
7.7 Additional requirements for start-up procedure	17
7.8 Additional requirements for operating instructions	17
7.8.1 General	17
7.8.2 Lay operator operating instructions	17
7.8.3 Healthcare professional operator operating instructions	18
7.9 Cleaning, disinfection, and sterilization	18
7.10 Additional requirements for maintenance	18
7.11 Additional requirements for accessories, supplementary equipment, used material	18
7.12 Additional requirements for the technical description	18
8 Protection against electrical hazards from ME equipment	19
9 Protection against mechanical hazards of ME equipment and ME systems	19
9.1 General	19
9.2 Additional requirements for instability from unwanted lateral movement	19
9.3 Grips and other handling devices	19
10 Protection against unwanted and excessive radiation hazards	19
11 Protection against excessive temperatures and other hazards	19
11.1 General	19
11.2 Cleaning and disinfection of ME equipment or ME system	20
11.3 Sterilization of ME equipment or ME system	20

11.4	<i>Biocompatibility of ME equipment and ME systems</i>	20
11.5	<i>Interruption of the power supply / supply mains to ME equipment</i>	21
11.5.1	General.....	21
11.5.2	Power sources.....	21
11.5.3	Alternative power supply/supply mains.....	22
12	Accuracy of controls and instruments and protection against hazardous outputs	22
12.1	General.....	22
12.2	Accuracy of controls and instruments.....	22
12.3	Accuracy of controls and instruments.....	23
12.3.1	General.....	23
12.3.2	Cardiorespiratory monitoring.....	23
12.3.3	Direct monitoring - respiration.....	23
12.3.4	Indirect monitoring – heart rate.....	23
12.3.5	Indirect monitoring from pulse oximetry.....	23
12.3.6	Apnoeic patient alarm condition.....	24
12.3.7	Sensor fault.....	25
12.3.8	Clinical performance evaluation.....	25
12.4	Usability of ME equipment.....	25
13	Hazardous situations and fault conditions for ME equipment	26
14	Programmable electrical medical systems (PEMS)	26
15	Construction of ME equipment	26
15.1	General.....	26
15.2	Mode of operation.....	26
15.3	Pre-use check.....	26
16	ME systems	27
17	Electromagnetic compatibility of ME equipment and ME systems	27
18	Requirements for the accessories	27
18.1	General.....	27
18.2	Labelling.....	27
19	Training	27
20	Functional connection	27
20.1	General.....	27
20.2	Connection to an electronic health record.....	27
20.3	Connection to a distributed alarm system.....	27
21	Electromagnetic disturbances – Requirements and tests	28
21.1	General.....	28
21.2	Compliance criteria.....	28
21.3	Requirements applicable to all ME equipment and ME systems.....	28
21.4	Additional general requirements.....	28
22	Usability	28
22.1	General.....	28
22.2	Primary operating functions.....	29
23	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	29
23.1	General.....	29
23.2	Volume and characteristics of auditory alarm signals and information signals.....	29
23.3	Additional requirements for termination of alarm signal inactivation.....	29
23.4	Additional requirements for alarm system logging.....	30
24	Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	30
Annex A (informative) General guidance and rationale		31
Annex B (informative) Sequence of testing		36

Annex C (informative) Guide to <i>marking</i> and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>	37
Annex D (informative) <i>Symbols on marking</i>	40
Annex E (informative) Examples of the connection of the measuring device (MD) for measurement of the <i>patient leakage current</i> and <i>patient auxiliary current</i>	41
Annex F (informative) Suitable measuring supply circuits	42
Annex G (informative) Protection against <i>hazards</i> of ignition of flammable anaesthetic mixtures	43
Annex H (informative) <i>PEMS</i> structure, <i>PEMS development life-cycle</i> and documentation	44
Annex I (informative) <i>ME systems</i> aspects	45
Annex J (informative) Survey of insulation paths	46
Annex K (informative) Simplified <i>patient leakage current</i> diagrams	47
Annex L (informative) Insulated winding wires for use without interleaved insulation	48
Annex M (informative) Reduction of pollution degrees	49
Annex N (informative) Data interface requirements	50
Annex O (informative) Considerations for a clinical performance study	53
Annex P (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances	55
Annex Q (informative) Reference to the <i>essential principles</i>	57
Annex R (informative) Reference to the general safety and performance requirements	59
Bibliography	62
Terminology — Alphabetized index of defined terms	63