

ISO 11608-4:2022-04 (E)

Needle-based injection systems for medical use - Requirements and test methods - Part 4: Needle-based injection systems containing electronics

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms and definitions	3
4	Abbreviated terms	11
5	General requirements	11
5.1	Conditions for application of NIS-E	11
5.2	General design requirements	12
5.3	Risk approach and usability engineering	12
6	General requirements for testing	13
6.1	Type tests	13
6.2	Number of samples	13
6.3	Ambient temperature, humidity, atmospheric pressure	18
7	Identification and marking of NIS-E	18
8	Protection against electrical hazards	18
8.1	General	18
8.2	Humidity preconditioning treatment	18
8.3	Requirements and test methods	19
8.3.1	General	19
8.3.2	Applied parts	19
8.3.3	Requirements related to power sources	21
8.3.4	Limitation of current for accessible parts and applied parts	22
8.4	Separation of parts (Type X and Type Y)	22
8.4.1	Means of protection (MOP)	22
8.4.2	Separation of patient connection	23
8.4.3	Maximum mains voltage	24
8.4.4	Working voltage	24
8.5	Patient leakage current and touch current (Type X and Type Y NIS-E)	25
8.5.1	General	25
8.5.2	Measurement of patient leakage current	29
8.5.3	Measurement of touch current	32
8.6	Insulation (Type X and Type Y)	33
8.6.1	General	33
8.6.2	Distance through solid insulation or use of thin sheet material	33
8.6.3	Dielectric strength	34
8.7	Insulation other than wire insulation	34
8.7.1	Mechanical strength and resistance to heat	34
8.8	Creepage distances and air clearances (Type X and Type Y NIS-E)	35
8.8.1	General	35
8.9	Specific hazardous situations	36
8.9.1	General	36

8.9.2	Emissions, deformation of enclosure or exceeding maximum temperature	36
8.9.3	Exceeding leakage current or voltage limits	38
8.9.4	Specific mechanical hazards	38
8.10	Single fault conditions (Type X and Type Y)	38
8.10.1	General	38
8.10.2	Failure of thermostats and temperature limiting devices	38
8.10.3	Leakage of liquid from batteries	39
8.10.4	Locking of moving parts	39
8.10.5	Additional test criteria for motor-operated NIS-E	39
ISO 11608-4:2022(E) 8.10.6 NIS-E intended for used in conjunction with oxygen rich environments		39
8.10.7	Power supply (Type Y)	39
8.11	Pre-conditioning for the influence of fluid leakage	40
9	Electromagnetic compatibility (EMC)	41
9.1	General requirements	41
9.1.1	Risk approach process for NIS-E	41
9.1.2	Non-medical electrical equipment used with NIS-E	41
9.1.3	General test conditions	42
9.2	NIS-E identification, marking and documents	47
9.2.1	Instruction for use in relation to EMC	47
9.2.2	Documentation of the tests	47
9.3	Electromagnetic emissions requirements for NIS-E	48
9.3.1	Protection of radio services and other equipment	48
9.3.2	Protection of the public mains network	48
9.3.3	Emissions requirements summary (Type X and Type Y)	49
9.4	Electromagnetic immunity requirements for NIS-E	49
9.4.1	General	49
9.4.2	Operating modes	51
9.4.3	Non-medical electrical equipment	51
9.4.4	Immunity test levels	51
9.4.5	Immunity to proximity fields from RF wireless communications equipment	56
9.4.6	Immunity to proximity magnetic fields in the frequency range 9 kHz to 13,56 MHz	58
10	Protection against mechanical hazards	58
10.1	General	58
10.2	Shock	58
10.3	Vibration	58
10.3.1	Sinusoidal vibration	58
10.3.2	Random vibration	58
10.4	Impact of OBDS enclosures	59
10.5	Push	59
11	Programmable NIS-E	59
Annex A (informative) Identification of immunity pass/fail criteria		60
Annex B (informative) Rationale for using 240 V for testing some requirements		62
Bibliography		63